



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB122858

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	11/11/2022	17:15
2	CAL2	1	Water	NA	NA	20.7	7.00	11/11/2022	17:16
3	CAL3	1	Water	NA	NA	20.7	10.03	11/11/2022	17:18
4	ICV	1	Water	NA	NA	20.8	6.98	11/11/2022	17:20
5	CCV1	1	Water	NA	NA	20.8	2.01	11/11/2022	17:22
6	N5561-02	1	Solid	20.02	20	20.4	7.67	11/11/2022	17:25
7	N5561-04	1	Solid	20.03	20	20.2	8.94	11/11/2022	17:26
8	N5561-06	1	Solid	20.04	20	20.5	7.79	11/11/2022	17:30
9	N5598-02	1	Solid	20.02	20	21.6	8.46	11/11/2022	17:32
10	N5598-02DUP	1	Solid	20.03	20	21.7	8.47	11/11/2022	17:33
11	N5607-02	1	Solid	20.04	20	21.7	8.58	11/11/2022	17:35
12	N5607-04	1	Solid	20.02	20	21.7	7.40	11/11/2022	17:40
13	N5607-06	1	Solid	20.03	20	21.7	7.80	11/11/2022	17:42
14	N5607-08	1	Solid	20.05	20	21.6	10.45	11/11/2022	17:49
15	N5607-10	1	Solid	20.02	20	21.8	8.70	11/11/2022	17:50
16	CCV2	1	Water	NA	NA	20.7	12.03	11/11/2022	17:52
17	N5607-12	1	Solid	20.03	20	21.9	8.77	11/11/2022	17:55
18	CCV3	1	Water	NA	NA	20.7	2.01	11/11/2022	18:00

WP 122858

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity n5706

WorkList ID : 164956

Department : Wet-Chemistry

Date : 11-11-2022 16:53:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/09/2022	Solid	N5561-02	Corrosivity	Cool 4 deg C	loui01	G11	WC-25	11/09/2022	9045D
11/09/2022	Solid	N5561-04	Corrosivity	Cool 4 deg C	loui01	G11	WC-06	11/09/2022	9045D
11/09/2022	Solid	N5561-06	Corrosivity	Cool 4 deg C	loui01	G11	WC-05	11/09/2022	9045D
11/11/2022	Solid	N5598-02	Corrosivity	Cool 4 deg C	ICES02	G11	FRAC-TANK-0760533	11/11/2022	9045D
11/10/2022	Solid	N5607-02	Corrosivity	Cool 4 deg C	loui01	G11	WC-01	11/10/2022	9045D
11/10/2022	Solid	N5607-04	Corrosivity	Cool 4 deg C	loui01	G11	WC-36	11/10/2022	9045D
11/10/2022	Solid	N5607-06	Corrosivity	Cool 4 deg C	loui01	G11	WC-37	11/10/2022	9045D
11/10/2022	Solid	N5607-08	Corrosivity	Cool 4 deg C	loui01	G11	WC-38	11/10/2022	9045D
11/10/2022	Solid	N5607-10	Corrosivity	Cool 4 deg C	loui01	G11	WC-39	11/10/2022	9045D
11/10/2022	Solid	N5607-12	Corrosivity	Cool 4 deg C	loui01	G11	WC-40	11/10/2022	9045D

Date/Time 11-11-22 17:10

Raw Sample Received by: J0 (WC)

Raw Sample Relinquished by: BH (WC)

Date/Time 11-11-22

Raw Sample Received by: BH (WC)

Raw Sample Relinquished by: J0 (WC)

18 + 30