



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB122882

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	11/14/2022	17:10
2	CAL2	1	Water	NA	NA	20.7	7.00	11/14/2022	17:12
3	CAL3	1	Water	NA	NA	20.8	10.03	11/14/2022	17:13
4	ICV	1	Water	NA	NA	20.8	6.99	11/14/2022	17:15
5	CCV1	1	Water	NA	NA	20.8	2.01	11/14/2022	17:16
6	N5605-01	1	Water	NA	NA	21.2	11.90	11/14/2022	17:20
7	N5605-02	1	Water	NA	NA	21.2	11.99	11/14/2022	17:24
8	N5605-03	1	Water	NA	NA	21.2	11.58	11/14/2022	17:25
9	N5611-06	1	Water	NA	NA	20.7	7.54	11/14/2022	17:30
10	N5615-03	1	Water	NA	NA	20.7	7.45	11/14/2022	17:32
11	N5616-01	1	Water	NA	NA	20.1	5.84	11/14/2022	17:35
12	N5616-02	1	Water	NA	NA	20.8	6.12	11/14/2022	17:40
13	N5616-03	1	Water	NA	NA	20.4	5.65	11/14/2022	17:43
14	N5616-04	1	Water	NA	NA	20.4	5.92	11/14/2022	17:45
15	N5616-05	1	Water	NA	NA	20.4	5.87	11/14/2022	17:46
16	CCV2	1	Water	NA	NA	20.8	12.03	11/14/2022	17:50
17	N5616-06	1	Water	NA	NA	20.8	5.66	11/14/2022	17:53
18	N5616-07	1	Water	NA	NA	20.4	7.58	11/14/2022	17:56
19	N5616-07DUP	1	Water	NA	NA	20.5	7.59	11/14/2022	17:57
20	CCV3	1	Water	NA	NA	20.8	2.01	11/14/2022	18:00

13 122882

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N5615 W

WorkList ID : 165027

Department : Wet-Chemistry

Date : 11-14-2022 16:51:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/10/2022	Water	N5605-01	pH	Cool 4 deg C	ATCE02	G11	C10HPT-1	11/10/2022	9040C
11/10/2022	Water	N5605-02	pH	Cool 4 deg C	ATCE02	G11	C10HPT-2	11/10/2022	9040C
11/10/2022	Water	N5605-03	pH	Cool 4 deg C	ATCE02	G11	C10HPT-3	11/10/2022	9040C
11/14/2022	Water	N5611-06	pH	Cool 4 deg C	PSEG03	G11	34081	11/14/2022	9040C
11/14/2022	Water	N5615-03	pH	Cool 4 deg C	PSEG03	G11	33575	11/14/2022	9040C
11/14/2022	Water	N5616-01	pH	Cool 4 deg C	PSEG03	K11	34120	11/14/2022	9040C
11/14/2022	Water	N5616-02	pH	Cool 4 deg C	PSEG03	K11	34121	11/14/2022	9040C
11/14/2022	Water	N5616-03	pH	Cool 4 deg C	PSEG03	K11	34108	11/14/2022	9040C
11/14/2022	Water	N5616-04	pH	Cool 4 deg C	PSEG03	K11	34109	11/14/2022	9040C
11/14/2022	Water	N5616-05	pH	Cool 4 deg C	PSEG03	K11	34110	11/14/2022	9040C
11/14/2022	Water	N5616-06	pH	Cool 4 deg C	PSEG03	K11	34111	11/14/2022	9040C
11/14/2022	Water	N5616-07	pH	Cool 4 deg C	PSEG03	K11	34112	11/14/2022	9040C

Date/Time 11/14/22 17:00
 Raw Sample Received by: 70 (WCC) (GDC)
 Raw Sample Relinquished by: 70 (WCC) (GDC)

Date/Time 11/14/22 18:30
 Raw Sample Received by: 70 (WCC) (GDC)
 Raw Sample Relinquished by: 70 (WCC) (GDC)