



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB122888

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	11/15/2022	08:45
2	N5611-06		1	12	100.00	>212.0	>212.0	11/15/2022	09:15
3	N5611-06DUP		1	11	100.00	>212.0	>212.0	11/15/2022	09:45
4	N5615-03		1	14	100.00	>212.0	>212.0	11/15/2022	10:15
5	N5616-01		1	13	100.00	>212.0	>212.0	11/15/2022	10:45
6	N5616-02		1	12	100.00	>212.0	>212.0	11/15/2022	11:15
7	N5616-03		1	13	100.00	>212.0	>212.0	11/15/2022	11:45
8	N5616-04		1	12	100.00	>212.0	>212.0	11/15/2022	12:15
9	N5616-05		1	13	100.00	>212.0	>212.0	11/15/2022	12:45
10	N5616-06		1	14	100.00	>212.0	>212.0	11/15/2022	13:15
11	N5616-07		1	12	100.00	>212.0	>212.0	11/15/2022	13:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16122888

WorkList Name: fp11-14

WorkList ID: 164998

Department: Wet-Chemistry

Date: 11-14-2022 08:38:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/21/2022	Water	N5611-06	Flash Point	Cool 4 deg C	PSEG03	G11	34081	11/14/2022	1010B
11/21/2022	Water	N5615-03	Flash Point	Cool 4 deg C	PSEG03	G11	33575	11/14/2022	1010B
11/21/2022	Water	N5616-01	Flash Point	Cool 4 deg C	PSEG03	K11	34120	11/14/2022	1010B
11/21/2022	Water	N5616-02	Flash Point	Cool 4 deg C	PSEG03	K11	34121	11/14/2022	1010B
11/21/2022	Water	N5616-03	Flash Point	Cool 4 deg C	PSEG03	K11	34108	11/14/2022	1010B
11/21/2022	Water	N5616-04	Flash Point	Cool 4 deg C	PSEG03	K11	34109	11/14/2022	1010B
11/21/2022	Water	N5616-05	Flash Point	Cool 4 deg C	PSEG03	K11	34110	11/14/2022	1010B
11/21/2022	Water	N5616-06	Flash Point	Cool 4 deg C	PSEG03	K11	34111	11/14/2022	1010B
11/21/2022	Water	N5616-07	Flash Point	Cool 4 deg C	PSEG03	K11	34112	11/14/2022	1010B

Date/Time 11/15/2022 08:40
 Raw Sample Received by: RA CWC
 Raw Sample Relinquished by: CP (SM)

Date/Time 11/15/2022 13:50
 Raw Sample Received by: CP (SM)
 Raw Sample Relinquished by: RA CWC