

LB122899


245.1

INSTRUMENT ID: CV4

Linear

 μ Abs.:

38004



A= 0.0000e+000

B= 2.6061e-004

C= 1.3583e-002

Rho= 0.9998011

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.021	0.021	29	0.000	29					
0.05	0.050	0.051	0.001	143	0.0 %	143	0				2
0.2	0.200	0.182	-0.018	646	0.0 %	646					-9
2.5	2.500	2.499	-0.001	9536	0.0 %	9536					0
5.0	5.000	4.922	-0.078	18836	0.0 %	18836					-2
7.5	7.500	7.657	0.157	29331	0.0 %	29331					2
10.0	10.000	9.918	-0.082	38004	0.0 %	38004					-1

LB122893 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 15 Nov 2022 12:42:33

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	29	-	0.0000	245.1	PPB	15 Nov 2022 13:47:12	S	Std
0.05 - 1	50.05	143	-	0.0500	245.1	PPB	15 Nov 2022 13:49:30	S	Std
0.2 - 1	50.2	646	-	0.2000	245.1	PPB	15 Nov 2022 13:51:47	S	Std
2.5 - 1	52.5	9536	-	2.5000	245.1	PPB	15 Nov 2022 13:54:04	S	Std
5.0 - 1	55	18836	-	5.0000	245.1	PPB	15 Nov 2022 13:56:21	S	Std
7.5 - 1	57.5	29331	-	7.5000	245.1	PPB	15 Nov 2022 13:58:37	S	Std
10.0 - 1	510	38004	-	10.0000	245.1	PPB	15 Nov 2022 14:08:35	S	Std
ICV22 - 1	ICV22	15112	3.9518	-	245.1	PPB	15 Nov 2022 14:13:53	U	SMPL
ICB22 - 1	ICB22	-43	0.0024	-	245.1	PPB	15 Nov 2022 14:16:08	U	SMPL
CCV33 - 1	CCV33	19482	5.0907	-	245.1	PPB	15 Nov 2022 14:18:23	U	SMPL
CCB33 - 1	CCB33	-88	-0.0094	-	245.1	PPB	15 Nov 2022 14:20:38	U	SMPL
CRA - 1	CRA	154	0.0537	-	245.1	PPB	15 Nov 2022 14:22:53	U	SMPL
HighStd - 1	HighStd	37655	9.8267	-	245.1	PPB	15 Nov 2022 14:25:09	U	SMPL
ChkStd - 1	ChkStd	27608	7.2084	-	245.1	PPB	15 Nov 2022 14:27:27	U	SMPL
PB149069BL - 1	PBW	-22	0.0078	-	245.1	PPB	15 Nov 2022 14:29:44	U	SMPL
N5534-01 - 1	B-3(OW)	-20	0.0084	-	245.1	PPB	15 Nov 2022 14:34:16	U	SMPL
N5564-01 - 1	B-4C(OW)	124	0.0459	-	245.1	PPB	15 Nov 2022 14:36:32	U	SMPL
N5564-01DUP - 1	B-4C(OW)DUP	48	0.0261	-	245.1	PPB	15 Nov 2022 14:38:49	U	SMPL
N5564-01MS - 1	B-4C(OW)MS	14301	3.7405	-	245.1	PPB	15 Nov 2022 14:41:06	U	SMPL
N5564-01MSD - 1	B-4C(OW)MSD	13277	3.4736	-	245.1	PPB	15 Nov 2022 14:43:21	U	SMPL
CCV34 - 1	CCV34	19336	5.0526	-	245.1	PPB	15 Nov 2022 14:45:36	U	SMPL
CCB34 - 1	CCB34	20	0.0188	-	245.1	PPB	15 Nov 2022 14:47:51	U	SMPL
N5564-02 - 1	B-1(OW)	37	0.0232	-	245.1	PPB	15 Nov 2022 14:50:07	U	SMPL
N5564-01LX5 - 1		77	0.0336	-	245.1	PPB	15 Nov 2022 14:52:22	U	SMPL
N5564-01A - 1		15574	4.0722	-	245.1	PPB	15 Nov 2022 14:54:37	U	SMPL
PB149069BS - 1	LCSW	14982	3.9180	-	245.1	PPB	15 Nov 2022 15:07:42	U	SMPL
CCV35 - 1	CCV35	19363	5.0597	-	245.1	PPB	15 Nov 2022 15:09:57	U	SMPL
CCB35 - 1	CCB35	27	0.0206	-	245.1	PPB	15 Nov 2022 15:12:14	U	SMPL