

LB122911


7471B

INSTRUMENT ID: CVL

Linear

μ Abs.:

45482



A= 0.0000e+000

B= 2.1701e-004

C= -7.9655e-003

Rho= 0.9997213

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Notes
0.0	0.000	-0.011	-0.011	-16	0.000	-16					
0.2	0.200	0.164	-0.036	791	0.0 %	791					-18
2.5	2.500	2.479	-0.021	11458	0.0 %	11458					-1
5.0	5.000	5.064	0.064	23370	0.0 %	23370					1
7.5	7.500	7.643	0.143	35258	0.0 %	35258					2
10.0	10.000	9.862	-0.138	45482	0.0 %	45482					-1

LB122911 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 16 Nov 2022 11:12:25

Sample ID	Extended ID	μ Abs	Conc	Stnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-16	-	0.0000	7471B	PPB	16 Nov 2022 11:14:29	S	Std
0.2 - 1	50.2	791	-	0.2000	7471B	PPB	16 Nov 2022 11:16:46	S	Std
2.5 - 1	52.5	11458	-	2.5000	7471B	PPB	16 Nov 2022 11:19:03	S	Std
5.0 - 1	55	23370	-	5.0000	7471B	PPB	16 Nov 2022 11:21:20	S	Std
7.5 - 1	57	35258	-	7.5000	7471B	PPB	16 Nov 2022 11:23:36	S	Std
10.0 - 1	51	45482	-	10.0000	7471B	PPB	16 Nov 2022 11:28:26	S	Std
ICV24 - 1	ICV24	19236	4.1665	-	7471B	PPB	16 Nov 2022 11:31:07	U	SMPL
ICB24 - 1	ICB24	12	-0.0054	-	7471B	PPB	16 Nov 2022 11:33:21	U	SMPL
CCV41 - 1	CCV41	23323	5.0534	-	7471B	PPB	16 Nov 2022 11:35:36	U	SMPL
CCB41 - 1	CCB41	-82	-0.0258	-	7471B	PPB	16 Nov 2022 11:37:52	U	SMPL
CRA - 1	CRA	783	0.1620	-	7471B	PPB	16 Nov 2022 11:40:07	U	SMPL
HighStd - 1	HighStd	45810	9.9333	-	7471B	PPB	16 Nov 2022 11:42:22	U	SMPL
ChkStd - 1	ChkStd	33532	7.2689	-	7471B	PPB	16 Nov 2022 11:44:37	U	SMPL
PB149111BL - 1	PBS	-39	-0.0164	-	7471B	PPB	16 Nov 2022 11:46:54	U	SMPL
PB149111BS - 1	LCSS	17155	3.7149	-	7471B	PPB	16 Nov 2022 11:49:10	U	SMPL
N5598-01 - 1	FRAC-TANK-0760533	657	0.1346	-	7471B	PPB	16 Nov 2022 11:51:26	U	SMPL
N5598-12 - 1	FRAC-TANK-SLURRY-0760533	-28	-0.0140	-	7471B	PPB	16 Nov 2022 11:53:43	U	SMPL
N5607-01 - 1	SB-01	7768	1.6778	-	7471B	PPB	16 Nov 2022 11:55:59	U	SMPL
N5607-03 - 1	SB-36	6355	1.3711	-	7471B	PPB	16 Nov 2022 11:58:17	U	SMPL
N5607-05 - 1	SB-37	8852	1.9130	-	7471B	PPB	16 Nov 2022 12:00:34	U	SMPL
CCV42 - 1	CCV42	23010	4.9855	-	7471B	PPB	16 Nov 2022 12:02:50	U	SMPL
CCB42 - 1	CCB42	-90	-0.0275	-	7471B	PPB	16 Nov 2022 12:05:05	U	SMPL
N5607-07 - 1	SB-38	1067	0.2236	-	7471B	PPB	16 Nov 2022 12:07:22	U	SMPL
N5607-09 - 1	SB-39	4594	0.9890	-	7471B	PPB	16 Nov 2022 12:09:36	U	SMPL
N5607-11 - 1	SB-40A	2711	0.5804	-	7471B	PPB	16 Nov 2022 12:11:52	U	SMPL
N5607-13 - 1	SB-40B	4723	1.0170	-	7471B	PPB	16 Nov 2022 12:14:08	U	SMPL
N5608-01 - 1	NP-1	376	0.0736	-	7471B	PPB	16 Nov 2022 12:16:24	U	SMPL
N5608-01DUP - 1	NP-1DUP	376	0.0736	-	7471B	PPB	16 Nov 2022 12:18:40	U	SMPL
N5608-01MS - 1	NP-1MS	18509	4.0087	-	7471B	PPB	16 Nov 2022 12:23:59	U	SMPL
N5608-01MSD - 1	NP-1MSD	18508	4.0081	-	7471B	PPB	16 Nov 2022 12:26:15	U	SMPL
N5613-01 - 1	WC-COMP-1	2566	0.5489	-	7471B	PPB	16 Nov 2022 12:28:32	U	SMPL
N5617-01 - 1	72-12013	836	0.1735	-	7471B	PPB	16 Nov 2022 12:30:47	U	SMPL
CCV43 - 1	CCV43	22583	4.8928	-	7471B	PPB	16 Nov 2022 12:33:02	U	SMPL
CCB43 - 1	CCB43	-78	-0.0249	-	7471B	PPB	16 Nov 2022 12:35:18	U	SMPL
N5619-01 - 1	OR-02-111422	3388	0.7273	-	7471B	PPB	16 Nov 2022 12:37:34	U	SMPL
N5608-01LX5 - 1		29	-0.0017	-	7471B	PPB	16 Nov 2022 12:39:49	U	SMPL
N5608-01A - 1		18294	3.9620	-	7471B	PPB	16 Nov 2022 12:42:05	U	SMPL
CCV44 - 1	CCV44	22472	4.8687	-	7471B	PPB	16 Nov 2022 12:44:21	U	SMPL
CCB44 - 1	CCB44	-79	-0.0251	-	7471B	PPB	16 Nov 2022 12:46:36	U	SMPL