

LB122916


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

41169



A= 0.0000e+000

B= 2.4082e-004

C= 1.5885e-002

Rho= 0.9997756

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.037	0.037	89	0.000	89					—
0.2	0.200	0.198	-0.002	756	0.0 %	756					-1
2.5	2.500	2.371	-0.129	9780	0.0 %	9780					-5
5.0	5.000	5.097	0.097	21098	0.0 %	21098					2
7.5	7.500	7.567	0.067	31354	0.0 %	31354					1
10.0	10.000	9.930	-0.070	41169	0.0 %	41169					-1

LB122916 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Nov 2022 14:46:02

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	89	-	0.0000	7470A	PPB	16 Nov 2022 14:48:10	S	Std
0.2 - 1	50.2	756	-	0.2000	7470A	PPB	16 Nov 2022 14:50:27	S	Std
2.5 - 1	52.5	9780	-	2.5000	7470A	PPB	16 Nov 2022 14:52:44	S	Std
5.0 - 1	55	21098	-	5.0000	7470A	PPB	16 Nov 2022 14:55:01	S	Std
7.5 - 1	57.5	31354	-	7.5000	7470A	PPB	16 Nov 2022 14:57:18	S	Std
10.0 - 1	510	41169	-	10.0000	7470A	PPB	16 Nov 2022 14:59:33	S	Std
ICV25 - 1	ICV25	15594	3.7713	-	7470A	PPB	16 Nov 2022 15:02:28	U	SMPL
ICB25 - 1	ICB25	90	0.0376	-	7470A	PPB	16 Nov 2022 15:04:43	U	SMPL
CCV45 - 1	CCV45	19645	4.7468	-	7470A	PPB	16 Nov 2022 15:06:58	U	SMPL
CCB45 - 1	CCB45	-69	-0.0007	-	7470A	PPB	16 Nov 2022 15:09:13	U	SMPL
CRA - 1	CRA	687	0.1813	-	7470A	PPB	16 Nov 2022 15:11:27	U	SMPL
HighStd - 1	HighStd	39869	9.6172	-	7470A	PPB	16 Nov 2022 15:13:42	U	SMPL
ChkStd - 1	ChkStd	26346	6.3606	-	7470A	PPB	16 Nov 2022 15:15:58	U	SMPL
PB149106BL - 1	PBW	-3	0.0152	-	7470A	PPB	16 Nov 2022 15:18:13	U	SMPL
PB149106BS - 1	LCSW	17174	4.1518	-	7470A	PPB	16 Nov 2022 15:20:30	U	SMPL
N5562-02 - 1	B-4C(OW)	2	0.0164	-	7470A	PPB	16 Nov 2022 15:22:47	U	SMPL
N5562-04 - 1	B-1(OW)	-15	0.0123	-	7470A	PPB	16 Nov 2022 15:25:03	U	SMPL
N5616-01 - 1	34120	361	0.1028	-	7470A	PPB	16 Nov 2022 15:27:20	U	SMPL
N5616-02 - 1	34121	62	0.0308	-	7470A	PPB	16 Nov 2022 15:29:37	U	SMPL
N5616-03 - 1	34108	633	0.1683	-	7470A	PPB	16 Nov 2022 15:32:58	U	SMPL
CCV46 - 1	CCV46	20701	5.0011	-	7470A	PPB	16 Nov 2022 15:35:13	U	SMPL
CCB46 - 1	CCB46	-61	0.0012	-	7470A	PPB	16 Nov 2022 15:37:30	U	SMPL
N5616-04 - 1	34109	1991	0.4954	-	7470A	PPB	16 Nov 2022 15:39:45	U	SMPL
N5616-05 - 1	34110	259	0.0783	-	7470A	PPB	16 Nov 2022 15:42:01	U	SMPL
N5616-06 - 1	34111	234	0.0722	-	7470A	PPB	16 Nov 2022 15:44:16	U	SMPL
N5616-06DUP - 1	34111DUP	277	0.0826	-	7470A	PPB	16 Nov 2022 15:46:32	U	SMPL
N5616-06MS - 1	34111MS	881	0.2280	-	7470A	PPB	16 Nov 2022 15:48:47	U	SMPL
N5616-06MSD - 1	34111MSD	901	0.2329	-	7470A	PPB	16 Nov 2022 15:51:04	U	SMPL
N5616-07 - 1	34112	1218	0.3092	-	7470A	PPB	16 Nov 2022 15:53:20	U	SMPL
N5616-06LX5 - 1		-82	-0.0039	-	7470A	PPB	16 Nov 2022 15:55:37	U	SMPL
N5616-06A - 1		12513	3.0293	-	7470A	PPB	16 Nov 2022 15:57:54	U	SMPL
CCV47 - 1	CCV47	20685	4.9973	-	7470A	PPB	16 Nov 2022 16:03:07	U	SMPL
CCB47 - 1	CCB47	32	0.0236	-	7470A	PPB	16 Nov 2022 16:05:22	U	SMPL