



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: jignesh

Run Number: LB122917

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP99867
calibration std. phosphate 0.5 ppm	WP99866
calibration std. phosphate 0.3 ppm	WP99865
calibration std. phosphate 0.1 ppm	WP99864
calibration std. phosphate 0.05 ppm	WP99863
calibration std. 0 ppm	WP99862
phosphate CCV std.	WP99868
Combined reagent	WP99926
Phenolphthalein indicator	WP96861
Sodium hydroxide, 1N	WP98644
Phosphate ICV-LCS Std	WP99869

Intercept: -0.0016

Slope: 0.6541

Regression: 0.999951

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	%D	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	0.002		11/16/2022	14:20
2	CAL2	0.05	1	50	50	0.033	0.053	-6	11/16/2022	14:20
3	CAL3	0.10	1	50	50	0.064	0.1	0	11/16/2022	14:21
4	CAL4	0.30	1	50	50	0.190	0.293	2.3	11/16/2022	14:21
5	CAL5	0.50	1	50	50	0.325	0.499	0.2	11/16/2022	14:22
6	CAL6	1.00	1	50	50	0.654	1.002	-0.2	11/16/2022	14:22



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: jignesh

Run Number: LB122917

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.318	0.489	11/16/2022	14:23
2	ICB		1	50	50	0.002	0.006	11/16/2022	14:23
3	CCV1	0.50	1	50	50	0.325	0.499	11/16/2022	14:24
4	CCB1		1	50	50	0.001	0.004	11/16/2022	14:24
5	PB149112BL		1	50	50	0.003	0.007	11/16/2022	14:25
6	PB149112BS	0.50	1	50	50	0.322	0.495	11/16/2022	14:25
7	N5546-01		1	50	50	1.030	1.577	11/16/2022	14:26
8	N5548-01		1	50	50	0.723	1.108	11/16/2022	14:26
9	N5548-01DUP		1	50	50	0.725	1.111	11/16/2022	14:27
10	N5548-01MS	0.50	1	50	50	0.988	1.513	11/16/2022	14:27
11	N5548-01MSD	0.50	1	50	50	0.986	1.510	11/16/2022	14:28
12	N5606-01		1	50	50	1.544	2.363	11/16/2022	14:28
13	N5546-01		10	50	50	0.241	0.371	11/16/2022	14:29
14	N5548-01		5	50	50	0.386	0.593	11/16/2022	14:29
15	CCV2	0.50	1	50	50	0.323	0.496	11/16/2022	14:30
16	CCB2		1	50	50	0.003	0.007	11/16/2022	14:30
17	N5548-01DUP		5	50	50	0.385	0.591	11/16/2022	14:31
18	N5548-01MS	0.50	5	50	50	0.437	0.671	11/16/2022	14:31
19	N5548-01MSD	0.50	5	50	50	0.438	0.672	11/16/2022	14:32
20	N5606-01		20	50	50	0.308	0.473	11/16/2022	14:32
21	CCV3	0.50	1	50	50	0.321	0.493	11/16/2022	14:33
22	CCB3		1	50	50	0.002	0.006	11/16/2022	14:33