



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: jignesh

Run Number: LB122918

BALANCE ID: WC SC-4

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP99867
calibration std. phosphate 0.5 ppm	WP99866
calibration std. phosphate 0.3 ppm	WP99865
calibration std. phosphate 0.1 ppm	WP99864
calibration std. phosphate 0.05 ppm	WP99863
calibration std. 0 ppm	WP99862
phosphate CCV std.	WP99868
Combined reagent	WP99926
Phenolphthalein indicator	WP96861
Sodium hydroxide, 1N	WP98644
Phosphate ICV-LCS Std	WP99869

Intercept: -0.0016

Slope: 0.6541

Regression: 0.999951

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	%D	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	0.002		11/16/2022	14:20
2	CAL2	0.05	1	50	50	0.033	0.053	-6	11/16/2022	14:20
3	CAL3	0.10	1	50	50	0.064	0.1	0	11/16/2022	14:21
4	CAL4	0.30	1	50	50	0.190	0.293	2.3	11/16/2022	14:21
5	CAL5	0.50	1	50	50	0.325	0.499	0.2	11/16/2022	14:22
6	CAL6	1.00	1	50	50	0.654	1.002	-0.2	11/16/2022	14:22



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: jignesh

Run Number: LB122918

BALANCE ID: WC SC-4

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.318	0.489	11/16/2022	14:23
2	ICB		1	50	50	0.002	0.006	11/16/2022	14:23
3	CCV1	0.50	1	50	50	0.326	0.501	11/16/2022	14:45
4	CCB1		1	50	50	0.003	0.007	11/16/2022	14:45
5	PB149113BL		1	1.00	50	0.002	0.006	11/16/2022	14:46
6	PB149113BS	0.50	1	1.00	50	0.323	0.496	11/16/2022	14:46
7	N5516-04		1	1.03	50	2.241	3.429	11/16/2022	14:47
8	N5516-04DUP		1	1.01	50	2.246	3.436	11/16/2022	14:47
9	N5516-04MS	0.50	1	1.05	50	2.256	3.451	11/16/2022	14:48
10	N5516-04MSD	0.50	1	1.05	50	2.257	3.453	11/16/2022	14:48
11	N5516-08		1	1.06	50	2.142	3.277	11/16/2022	14:49
12	N5516-04		100	1.03	50	0.621	0.952	11/16/2022	14:49
13	N5516-04DUP		100	1.01	50	0.621	0.952	11/16/2022	14:50
14	N5516-04MS	0.50	100	1.05	50	0.652	0.999	11/16/2022	14:50
15	CCV2	0.50	1	50	50	0.318	0.489	11/16/2022	14:51
16	CCB2		1	50	50	0.003	0.007	11/16/2022	14:51
17	N5516-04MSD	0.50	100	1.05	50	0.651	0.998	11/16/2022	14:52
18	N5516-08		100	1.06	50	0.435	0.667	11/16/2022	14:53
19	CCV3	0.50	1	50	50	0.324	0.498	11/16/2022	14:53
20	CCB3		1	50	50	0.002	0.006	11/16/2022	14:54