



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB122925

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	11/17/2022	10:30
2	N5650-01		1	12	100.00	>212.0	>212.0	11/17/2022	11:00
3	N5650-01DUP		1	14	100.00	>212.0	>212.0	11/17/2022	11:30
4	N5650-02		1	14	100.00	>212.0	>212.0	11/17/2022	12:00
5	N5650-03		1	13	100.00	>212.0	>212.0	11/17/2022	12:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

L6122925

WorkList Name : fp11-16

WorkList ID : 165074

Department : Wet-Chemistry

Date : 11-16-2022 08:45:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/23/2022	Water	N5650-01	Flash Point	Cool 4 deg C	PSEG03	G11	34270	11/16/2022	1010B
11/23/2022	Water	N5650-02	Flash Point	Cool 4 deg C	PSEG03	G11	34271	11/16/2022	1010B
11/23/2022	Water	N5650-03	Flash Point	Cool 4 deg C	PSEG03	G11	34267	11/16/2022	1010B

Date/Time 11/17/22 09:45
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: CP(SM)

Date/Time 11/17/22 12:40
 Raw Sample Received by: CP(SM)
 Raw Sample Relinquished by: RM CWC