



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB122926

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	11/17/2022	10:25
2	CAL2	1	Water	NA	NA	20.9	7.00	11/17/2022	10:26
3	CAL3	1	Water	NA	NA	20.9	10.03	11/17/2022	10:27
4	ICV	1	Water	NA	NA	20.8	6.98	11/17/2022	10:30
5	CCV1	1	Water	NA	NA	20.8	2.01	11/17/2022	10:32
6	N5650-01	1	Water	NA	NA	20.9	6.81	11/17/2022	10:35
7	N5650-01DUP	1	Water	NA	NA	20.8	6.82	11/17/2022	10:36
8	N5650-02	1	Water	NA	NA	20.7	6.69	11/17/2022	10:39
9	N5650-03	1	Water	NA	NA	20.5	6.92	11/17/2022	10:40
10	N5675-02	1	Water	NA	NA	20.5	6.22	11/17/2022	10:42
11	N5675-04	1	Water	NA	NA	20.5	8.73	11/17/2022	10:45
12	N5675-06	1	Water	NA	NA	20.2	9.08	11/17/2022	10:48
13	N5675-07	1	Water	NA	NA	20.2	4.40	11/17/2022	10:50
14	N5675-08	1	Water	NA	NA	20.1	7.30	11/17/2022	10:52
15	CCV2	1	Water	NA	NA	20.8	12.03	11/17/2022	11:00

WORKLIST(Hardcopy Internal Chain)

122926

WorkList Name : ph n5675

WorkList ID : 165122

Department : Wet-Chemistry

Date : 11-17-2022 10:17:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/16/2022	Water	N5650-01	pH C	Cool 4 deg C	PSEG03	G11	34270	11/16/2022	9040C
11/16/2022	Water	N5650-02	pH C	Cool 4 deg C	PSEG03	G11	34271	11/16/2022	9040C
11/16/2022	Water	N5650-03	pH C	Cool 4 deg C	PSEG03	G11	34267	11/16/2022	9040C
11/16/2022	Water	N5675-02	pH A	Cool 4 deg C	PSEG03	K11	AUD-1914-TOTE	11/16/2022	9040C
11/16/2022	Water	N5675-04	pH A	Cool 4 deg C	PSEG03	K11	AUD-1742-TOTE	11/16/2022	9040C
11/16/2022	Water	N5675-06	pH A	Cool 4 deg C	PSEG03	K11	AUD-1912	11/16/2022	9040C
11/16/2022	Water	N5675-07	pH A	Cool 4 deg C	PSEG03	K11	AUD-1913	11/16/2022	9040C
11/16/2022	Water	N5675-08	pH A	Cool 4 deg C	PSEG03	K11	AUD-1245	11/16/2022	9040C

Date/Time 11-17-22 10:20

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 11-17-22

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]