



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB122929

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	11/17/2022	10:00
2	CAL2	1	Water	NA	NA	20.8	7.00	11/17/2022	10:01
3	CAL3	1	Water	NA	NA	20.8	10.03	11/17/2022	10:05
4	ICV	1	Water	NA	NA	20.7	6.99	11/17/2022	10:10
5	CCV1	1	Water	NA	NA	20.8	2.01	11/17/2022	10:15
6	N5644-01	1	Solid	20.02	20	20.9	6.72	11/17/2022	10:20
7	N5644-01DUP	1	Solid	20.01	20	20.8	6.73	11/17/2022	10:21
8	N5644-02	1	Solid	20.04	20	20.8	9.86	11/17/2022	10:27
9	CCV2	1	Water	NA	NA	20.6	12.03	11/17/2022	10:30

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n5644

WorkList ID : 165123

Department : Wet-Chemistry

Date : 11-17-2022 10:18:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/16/2022	Solid	N5644-01	pH	Cool 4 deg C	PSEG03	K11	102122	11/16/2022	9045D
11/16/2022	Solid	N5644-02	pH	Cool 4 deg C	PSEG03	K11	110222	11/16/2022	9045D

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

11-17-22 10:25

90 (WC)

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

11-17-22

12:40

90

90