

Analysis Method: 9034

Parameter: Reactive Sulfide

Run Number: LB122930

ANALYST: Iwona

SUPERVISOR REVIEW BY: Rubina

Constant: 16000

Normality1: 0.025

Normality2: 0.025

| Reagent/Standard                   | Lot/Log # |
|------------------------------------|-----------|
| SODIUM THIOSULFATE, 0.025N, 4LITRE | W2941     |
| IODINE SOLUTION .025N 1L           | W2924     |
| Starch Solution, 4L                | W2977     |

| Seq | Lab ID      | True Value (mg/l) | DF | Initial Weight (g) | Final Volume (ml) | T1 (ml) | T2 Initial | T2 Final | T2 Diff. (ml) | T1 - T2 Diff (mL) | Value Corrected With Blank | Result (ppm) | Anal Date  | Anal Time |
|-----|-------------|-------------------|----|--------------------|-------------------|---------|------------|----------|---------------|-------------------|----------------------------|--------------|------------|-----------|
| 1   | PB149116BL  |                   | 1  | 5.00               | 50                | 2.00    | 0.00       | 1.92     | 1.92          | 0.08              | 0.00                       | 0.00         | 11/17/2022 | 09:30     |
| 2   | N5607-02    |                   | 1  | 5.04               | 50                | 2.00    | 0.00       | 1.86     | 1.86          | 0.14              | 0.06                       | 4.76         | 11/17/2022 | 09:32     |
| 3   | N5607-02DUP |                   | 1  | 5.02               | 50                | 2.00    | 0.00       | 1.82     | 1.82          | 0.18              | 0.10                       | 7.97         | 11/17/2022 | 09:35     |
| 4   | N5607-04    |                   | 1  | 5.02               | 50                | 2.00    | 0.00       | 1.88     | 1.88          | 0.12              | 0.04                       | 3.19         | 11/17/2022 | 09:37     |
| 5   | N5607-06    |                   | 1  | 5.06               | 50                | 2.00    | 0.00       | 1.90     | 1.90          | 0.10              | 0.02                       | 1.58         | 11/17/2022 | 09:40     |
| 6   | N5607-08    |                   | 1  | 5.02               | 50                | 2.00    | 0.00       | 1.86     | 1.86          | 0.14              | 0.06                       | 4.78         | 11/17/2022 | 09:43     |
| 7   | N5607-10    |                   | 1  | 5.05               | 50                | 2.00    | 0.00       | 1.84     | 1.84          | 0.16              | 0.08                       | 6.34         | 11/17/2022 | 09:45     |
| 8   | N5607-12    |                   | 1  | 5.03               | 50                | 2.00    | 0.00       | 1.88     | 1.88          | 0.12              | 0.04                       | 3.18         | 11/17/2022 | 09:48     |
| 9   | N5613-02    |                   | 1  | 5.07               | 50                | 2.00    | 0.00       | 1.90     | 1.90          | 0.10              | 0.02                       | 1.58         | 11/17/2022 | 09:50     |
| 10  | N5617-03    |                   | 1  | 5.04               | 50                | 2.00    | 0.00       | 1.90     | 1.90          | 0.10              | 0.02                       | 1.59         | 11/17/2022 | 09:53     |

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 \* Normality1) - ((T1 - Value Corrected With Blank) \* Normality2)) \* Constant / Initial Volume