


LB122932

SFAM01.1

INSTRUMENT INDICV

 μ Abs.:

40730



Linear

A= 0.0000e+000

B= 2.4250e-004 *slope*C= 1.8952e-002 *= 1.8952e-002*

Rho= 0.9996404

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.042	0.042	94	0.000	94					
0.05	0.050					185	0				
0.20	0.200	0.215	0.015	810	0.0 %	810					
2.50	2.500	2.379	-0.121	9733	0.0 %	9733					
5.00	5.000	4.991	-0.009	20503	0.0 %	20503					
7.50	7.500	7.677	0.177	31579	0.0 %	31579					
10.0	10.000	9.896	-0.104	40730	0.0 %	40730					

11/22/22

LB122932

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	94	-		0 SFAM01.1	PPB	11/17/2022 12:45	Std
	0.2 S01	810	-		0.2 SFAM01.1	PPB	11/17/2022 12:48	Std
	2.5 S02	9733	-		2.5 SFAM01.1	PPB	11/17/2022 12:50	Std
	5 S03	20503	-		5 SFAM01.1	PPB	11/17/2022 12:52	Std
	7.5 S04	31579	-		7.5 SFAM01.1	PPB	11/17/2022 12:55	Std
	10 S05	40730	-		10 SFAM01.1	PPB	11/17/2022 12:57	Std
ICV026	ICV026	16546	4.0313 -		SFAM01.1	PPB	11/17/2022 13:00	SMPL
ICB026	ICB026	176	0.0616 -		SFAM01.1	PPB	11/17/2022 13:02	SMPL
CCV048	CCV048	20772	5.0561 -		SFAM01.1	PPB	11/17/2022 13:04	SMPL
CCB048	CCB048	21	0.024 -		SFAM01.1	PPB	11/17/2022 13:06	SMPL
PB149139BL	PBW139	173	0.0609 -		SFAM01.1	PPB	11/17/2022 13:09	SMPL
PB149107TB	LEB139	147	0.0546 -		SFAM01.1	PPB	11/17/2022 13:11	SMPL
N5403-01	GBM92	307	0.0934 -		SFAM01.1	PPB	11/17/2022 13:13	SMPL
N5403-02	GBM93	163	0.0585 -		SFAM01.1	PPB	11/17/2022 13:15	SMPL
N5403-03	GBM93D	160	0.0578 -		SFAM01.1	PPB	11/17/2022 13:18	SMPL
N5403-04	GBM93S	4523	1.1158 -		SFAM01.1	PPB	11/17/2022 13:20	SMPL
N5403-05	GBM94	574	0.1581 -		SFAM01.1	PPB	11/17/2022 13:22	SMPL
N5403-06	GBM95	536	0.1489 -		SFAM01.1	PPB	11/17/2022 13:24	SMPL
N5403-07	GBM96	586	0.1611 -		SFAM01.1	PPB	11/17/2022 13:27	SMPL
N5403-08	GBM97	393	0.1143 -		SFAM01.1	PPB	11/17/2022 13:29	SMPL
N5403-09	GBM98	98	0.0427 -		SFAM01.1	PPB	11/17/2022 13:31	SMPL
N5403-10	GBM99	135	0.0517 -		SFAM01.1	PPB	11/17/2022 13:34	SMPL
N5403-11	GBMA0	109	0.0454 -		SFAM01.1	PPB	11/17/2022 13:36	SMPL
N5403-12	GBMA1	117	0.0473 -		SFAM01.1	PPB	11/17/2022 13:38	SMPL
CCV049	CCV049	19762	4.8112 -		SFAM01.1	PPB	11/17/2022 13:40	SMPL
CCB049	CCB049	52	0.0316 -		SFAM01.1	PPB	11/17/2022 13:43	SMPL