

LB122950


SFAM01.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

52987



A= 0.0000e+000

B= 1.8899e-004 slope

C= -5.6933e-002 *intercept*

Rho= 0.9997241

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.032	0.032	469	0.000	469					
0.05	0.050					185	0				
0.20	0.200	0.218	0.018	1457	0.0 %	1457					
2.50	2.500	2.483	-0.017	13437	0.0 %	13437					
5.00	5.000	4.861	-0.139	26023	0.0 %	26023					
7.50	7.500	7.649	0.149	40776	0.0 %	40776					
10.0	10.000	9.957	-0.043	52987	0.0 %	52987					

11-9-11
022010

LB122950
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	469	-		0 SFAM01.1	PPB	11/18/2022 15:08	Std
	0.2 S01	1457	-		0.2 SFAM01.1	PPB	11/18/2022 15:10	Std
	2.5 S02	13437	-		2.5 SFAM01.1	PPB	11/18/2022 15:14	Std
	5 S03	26023	-		5 SFAM01.1	PPB	11/18/2022 15:16	Std
	7.5 S04	40776	-		7.5 SFAM01.1	PPB	11/18/2022 15:18	Std
	10 S05	52987	-		10 SFAM01.1	PPB	11/18/2022 15:21	Std
ICV028	ICV028	21670	4.0384 -		SFAM01.1	PPB	11/18/2022 15:23	SMPL
ICB028	ICB028	554	0.0478 -		SFAM01.1	PPB	11/18/2022 15:26	SMPL
CCV057	CCV057	26993	5.0444 -		SFAM01.1	PPB	11/18/2022 15:28	SMPL
CCB057	CCB057	483	0.0343 -		SFAM01.1	PPB	11/18/2022 15:30	SMPL
PB149181BL	PBS181	-30	-0.0626 -		SFAM01.1	PPB	11/18/2022 15:35	SMPL
N5714-01	BGC74	2332	0.3838 -		SFAM01.1	PPB	11/18/2022 15:37	SMPL
N5714-02	BGC74D	2588	0.4322 -		SFAM01.1	PPB	11/18/2022 15:39	SMPL
N5714-03	BGC74S	17387	3.229 -		SFAM01.1	PPB	11/18/2022 15:42	SMPL
N5714-04	BGC75	2129	0.3454 -		SFAM01.1	PPB	11/18/2022 15:44	SMPL
N5714-05	BGC76	1943	0.3103 -		SFAM01.1	PPB	11/18/2022 15:46	SMPL
N5714-06	BGC77	1681	0.2608 -		SFAM01.1	PPB	11/18/2022 15:48	SMPL
N5714-07	BGC78	2552	0.4254 -		SFAM01.1	PPB	11/18/2022 15:51	SMPL
N5714-08	BGC79	2560	0.4269 -		SFAM01.1	PPB	11/18/2022 15:53	SMPL
N5714-09	BGC80	3734	0.6487 -		SFAM01.1	PPB	11/18/2022 15:55	SMPL
CCV058	CCV058	26054	4.867 -		SFAM01.1	PPB	11/18/2022 16:03	SMPL
CCB058	CCB058	39	-0.0496 -		SFAM01.1	PPB	11/18/2022 16:05	SMPL