



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 11/21/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:30
In Date: 11/18/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:33
Out Date: 11/19/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122954

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N5720-01	EXG92	1	0.91	9.07	9.98	9.15	90.8	
N5720-02	EXG93	2	0.91	8.72	9.63	9.00	92.8	
N5720-03	EXG94	3	0.91	8.74	9.65	9.18	94.6	
N5720-04	EXG95	4	0.9	9.02	9.92	9.2	92.0	
N5720-05	EXG96	5	0.9	8.68	9.58	8.86	91.7	
N5720-06	EXG97	6	0.9	8.94	9.84	8.74	87.7	
N5720-07	EXG98	7	0.92	8.77	9.69	9.06	92.8	
N5720-08	EXG99	8	0.92	8.69	9.61	9.00	93.0	
N5720-09	EXGA0	9	0.92	8.70	9.62	8.94	92.2	
N5720-10	EXGA0MS	10	0.92	8.70	9.62	8.94	92.2	
N5720-11	EXGA0MSD	11	0.92	8.70	9.62	8.94	92.2	
N5720-12	EXGA1	12	0.91	8.91	9.82	9.15	92.5	
N5720-13	EXGA2	13	0.91	8.69	9.6	8.63	88.8	
N5720-14	EXGA3	14	0.89	8.95	9.84	8.59	86.0	
N5720-15	EXGA4	15	0.91	9.04	9.95	8.56	84.6	
N5720-16	EXGA5	16	0.91	8.77	9.68	8.85	90.5	
N5720-17	EXGA6	17	0.91	8.73	9.64	8.78	90.1	
N5720-18	EXGA7	18	0.91	9.03	9.94	9.12	90.9	
N5720-19	EXGA8	19	0.91	8.99	9.9	9.00	90.0	
N5720-20	EXGA9	20	0.9	9.02	9.92	7.71	75.5	
N5720-21	EXGB0	21	0.9	8.84	9.74	8.55	86.5	
N5720-22	EXGB1	22	0.89	8.80	9.69	8.95	91.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$