


74713
 11/21/27 SPAM01.1

LB122966
 INSTRUMENT ID: CVZ

Linear



A= 0.0000e+000
 B= 2.5828e-004
 C= -3.6429e-003
 Rho= 0.9998101
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.016	0.016	76	0.000	76					5
0.05	0.050					185	0				3
0.20	0.200	0.207	0.007	814	0.0 %	814					-1
2.50	2.500	2.479	-0.021	9612	0.0 %	9612					-2
5.00	5.000	4.922	-0.078	19072	0.0 %	19072					1
7.50	7.500	7.641	0.141	29600	0.0 %	29600					-1
10.0	10.000	9.935	-0.065	38480	0.0 %	38480					

LB122966 INSTRUMENT ID : CV1

Method: SFAM01.1

Operator: Admin

Date of Analysis: 21 Nov 2022 11:07:12

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.00 - 1	50	76	-	0.0000	SFAM01.1	PPB	21 Nov 2022 11:16:05	S	Std
0.20 - 1	50.2	814	-	0.2000	SFAM01.1	PPB	21 Nov 2022 11:18:23	S	Std
2.50 - 1	52.5	9612	-	2.5000	SFAM01.1	PPB	21 Nov 2022 11:20:41	S	Std
5.00 - 1	55	19072	-	5.0000	SFAM01.1	PPB	21 Nov 2022 11:22:58	S	Std
7.50 - 1	57.5	29600	-	7.5000	SFAM01.1	PPB	21 Nov 2022 11:25:14	S	Std
10.0 - 1	59	38480	-	10.0000	SFAM01.1	PPB	21 Nov 2022 11:30:11	S	Std
ICV29 - 1	ICV29	16685	4.3057	-	SFAM01.1	PPB	21 Nov 2022 11:33:03	U	SMPL
ICB29 - 1	ICB29	125	0.0286	-	SFAM01.1	PPB	21 Nov 2022 11:35:18	U	SMPL
CCV59 - 1	CCV59	20421	5.2706	-	SFAM01.1	PPB	21 Nov 2022 11:37:33	U	SMPL
CCB59 - 1	CCB59	-17	-0.0080	-	SFAM01.1	PPB	21 Nov 2022 11:39:48	U	SMPL
CRA - 1	CRA	816	0.2071	-	SFAM01.1	PPB	21 Nov 2022 11:42:03	U	SMPL
HighStd - 1	HighStd	38535	9.9491	-	SFAM01.1	PPB	21 Nov 2022 11:44:18	U	SMPL
ChkStd - 1	ChkStd	27518	7.1036	-	SFAM01.1	PPB	21 Nov 2022 11:46:33	U	SMPL
PB149184BL - 1	PBS	24	0.0026	-	SFAM01.1	PPB	21 Nov 2022 11:51:26	U	SMPL
PB149184BS - 1	LCSS	15473	3.9927	-	SFAM01.1	PPB	21 Nov 2022 11:53:41	U	SMPL
N5678-01 - 1	250FB-BP-BACKFILL-01	611	0.1542	-	SFAM01.1	PPB	21 Nov 2022 12:01:40	U	SMPL
N5678-01DUP - 1	250FB-BP-BACKFILL-01DUP	532	0.1338	-	SFAM01.1	PPB	21 Nov 2022 12:03:56	U	SMPL
N5678-01MS - 1	250FB-BP-BACKFILL-01MS	14566	3.7584	-	SFAM01.1	PPB	21 Nov 2022 12:06:13	U	SMPL
N5678-01MSD - 1	250FB-BP-BACKFILL-01MSD	13737	3.5443	-	SFAM01.1	PPB	21 Nov 2022 12:08:30	U	SMPL
N5678-03 - 1	250FB-BP-BACKFILL-02	677	0.1712	-	SFAM01.1	PPB	21 Nov 2022 12:10:45	U	SMPL
CCV60 - 1	CCV60	19254	4.9692	-	SFAM01.1	PPB	21 Nov 2022 12:13:00	U	SMPL
CCB60 - 1	CCB60	14	-0.0000	-	SFAM01.1	PPB	21 Nov 2022 12:15:15	U	SMPL
N5678-05 - 1	250FB-BP-BACKFILL-03	787	0.1996	-	SFAM01.1	PPB	21 Nov 2022 12:17:30	U	SMPL
N5678-07 - 1	250FB-BP-BACKFILL-04	678	0.1715	-	SFAM01.1	PPB	21 Nov 2022 12:19:46	U	SMPL
N5678-09 - 1	250FB-BP-BACKFILL-05	546	0.1374	-	SFAM01.1	PPB	21 Nov 2022 12:22:01	U	SMPL
N5678-11 - 1	250FB-BP-BACKFILL-06	716	0.1813	-	SFAM01.1	PPB	21 Nov 2022 12:24:18	U	SMPL
N5678-13 - 1	250FB-BP-BACKFILL-07	1090	0.2779	-	SFAM01.1	PPB	21 Nov 2022 12:26:34	U	SMPL
N5678-15 - 1	250FB-BP-BACKFILL-08	516	0.1296	-	SFAM01.1	PPB	21 Nov 2022 12:28:50	U	SMPL
N5678-17 - 1	250FB-BP-BACKFILL-09	534	0.1343	-	SFAM01.1	PPB	21 Nov 2022 12:31:06	U	SMPL
N5678-19 - 1	250FB-BP-BACKFILL-10	542	0.1363	-	SFAM01.1	PPB	21 Nov 2022 12:33:23	U	SMPL
N5678-21 - 1	250FB-BP-BACKFILL-11	618	0.1562	-	SFAM01.1	PPB	21 Nov 2022 12:35:40	U	SMPL
PB149185BL - 1	PBS	-70	-0.0217	-	SFAM01.1	PPB	21 Nov 2022 12:37:55	U	SMPL
CCV61 - 1	CCV61	18797	4.8512	-	SFAM01.1	PPB	21 Nov 2022 12:40:11	U	SMPL
CCB61 - 1	CCB61	20	0.0015	-	SFAM01.1	PPB	21 Nov 2022 12:42:26	U	SMPL
PB149185BS - 1	LCSS	13620	3.5141	-	SFAM01.1	PPB	21 Nov 2022 12:44:42	U	SMPL
N5678-01 - 1	ACD-DEBRIS-1117	618	0.1560	-	SFAM01.1	PPB	21 Nov 2022 12:46:57	U	SMPL
N5686-01 - 1	OK-2-111722	815	0.2069	-	SFAM01.1	PPB	21 Nov 2022 12:49:14	U	SMPL
N5691-01 - 1	ETGI-358	2858	0.7345	-	SFAM01.1	PPB	21 Nov 2022 12:56:37	U	SMPL
N5691-01DUP - 1	ETGI-358DUP	2934	0.7541	-	SFAM01.1	PPB	21 Nov 2022 13:01:38	U	SMPL
N5691-01MS - 1	ETGI-358MS	18198	4.6965	-	SFAM01.1	PPB	21 Nov 2022 13:12:53	U	SMPL
N5691-01MSD - 1	ETGI-358MSD	17864	4.6102	-	SFAM01.1	PPB	21 Nov 2022 13:15:09	U	SMPL
N5692-01 - 1	DF-SPILL-DRUM-11-3	801	0.2032	-	SFAM01.1	PPB	21 Nov 2022 13:17:26	U	SMPL
N5692-03 - 1	HS-SPILL-DRUM-11-16	388	0.0966	-	SFAM01.1	PPB	21 Nov 2022 13:19:42	U	SMPL
N5695-01 - 1	HD-1-111722	1766	0.4525	-	SFAM01.1	PPB	21 Nov 2022 13:21:58	U	SMPL
CCV62 - 1	CCV62	19414	5.0106	-	SFAM01.1	PPB	21 Nov 2022 13:24:14	U	SMPL
CCB62 - 1	CCB62	-14	-0.0073	-	SFAM01.1	PPB	21 Nov 2022 13:26:29	U	SMPL
N5696-01 - 1	EO-1-111722	4167	1.0726	-	SFAM01.1	PPB	21 Nov 2022 13:28:45	U	SMPL
N5697-01 - 1	SU-03-111722	1600	0.4096	-	SFAM01.1	PPB	21 Nov 2022 13:31:00	U	SMPL
N5698-01 - 1	BANK-RUN-SOIL-SAMPLE-1	518	0.1301	-	SFAM01.1	PPB	21 Nov 2022 13:33:16	U	SMPL
N5698-02 - 1	BANK-RUN-SOIL-SAMPLE-2	290	0.0713	-	SFAM01.1	PPB	21 Nov 2022 13:35:32	U	SMPL
N5698-03 - 1	BANK-RUN-SOIL-SAMPLE-3	256	0.0625	-	SFAM01.1	PPB	21 Nov 2022 13:37:48	U	SMPL
N5678-01LX5 - 1		42	0.0072	-	SFAM01.1	PPB	21 Nov 2022 13:40:04	U	SMPL
N5678-01A - 1		16171	4.1730	-	SFAM01.1	PPB	21 Nov 2022 13:42:21	U	SMPL
N5691-01LX5 - 1		683	0.1728	-	SFAM01.1	PPB	21 Nov 2022 13:44:37	U	SMPL
N5691-01A - 1		17012	4.3902	-	SFAM01.1	PPB	21 Nov 2022 13:46:54	U	SMPL
CCV63 - 1	CCV63	18362	4.7389	-	SFAM01.1	PPB	21 Nov 2022 13:49:10	U	SMPL
CCB63 - 1	CCB63	49	0.0090	-	SFAM01.1	PPB	21 Nov 2022 13:51:26	U	SMPL

03
11/21/22