



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB122969

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	11/21/2022	12:45
2	CAL2	1	Water	NA	NA	20.7	7.00	11/21/2022	12:46
3	CAL3	1	Water	NA	NA	20.7	10.03	11/21/2022	12:47
4	ICV	1	Water	NA	NA	20.8	6.99	11/21/2022	12:50
5	CCV1	1	Water	NA	NA	20.8	2.01	11/21/2022	13:25
6	N5734-02	1	Water	NA	NA	20.3	6.16	11/21/2022	13:26
7	N5734-02DUP	1	Water	NA	NA	20.4	6.17	11/21/2022	13:27
8	CCV2	1	Water	NA	NA	20.8	12.03	11/21/2022	13:30

WORKLIST(Hardcopy Internal Chain)

VB 122969

WorkList Name : corrosivity n5740 w **WorkList ID :** 165238 **Department :** Wet-Chemistry **Date :** 11-21-2022 12:06:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/18/2022	Water	N5734-02	Corrosivity	Cool 4 deg C	PSEG03	K11	COMP-1	11/18/2022	9040C

Date/Time 11-21-22 12:10
 Raw Sample Received by: JD CSM
 Raw Sample Relinquished by: JD CSM

Date/Time 11-21-22 14:00
 Raw Sample Received by: JD CSM
 Raw Sample Relinquished by: JD CSM