



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB122970

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	11/21/2022	12:45
2	CAL2	1	Water	NA	NA	20.7	7.00	11/21/2022	12:46
3	CAL3	1	Water	NA	NA	20.7	10.03	11/21/2022	12:47
4	ICV	1	Water	NA	NA	20.8	6.98	11/21/2022	12:50
5	CCV1	1	Water	NA	NA	20.6	2.01	11/21/2022	13:45
6	N5709-01	1	Water	NA	NA	20.1	7.52	11/21/2022	13:46
7	N5728-01	1	Water	NA	NA	20.1	4.58	11/21/2022	13:49
8	N5732-01	1	Water	NA	NA	20.9	9.45	11/21/2022	14:00
9	N5732-01DUP	1	Water	NA	NA	20.8	9.46	11/21/2022	14:01
10	CCV2	1	Water	NA	NA	20.7	12.03	11/21/2022	14:10

WORKLIST(Hardcopy Internal Chain)

VB 122980

WorkList Name : ph n5732

WorkList ID : 165232

Department : Wet-Chemistry

Date : 11-21-2022 11:26:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/17/2022	Water	N5709-01	pH	Cool 4 deg C	ENV170	N11	COMP-SAMPLE-TANK-12-31-32-33	11/17/2022	9040C
11/17/2022	Water	N5728-01	pH	Cool 4 deg C	STER05	N11	POOL-WATER	11/17/2022	9040C
11/17/2022	Water	N5732-01	pH	Cool 4 deg C	ATCE02	K11	9DMC986SP	11/17/2022	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

11.21.22 11:30

JP (cc)

JP (cc)

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

11.21.22

JP (cc)

JP (cc)