



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB122972

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	11/21/2022	13:50
2	N5732-01		1	12	100.00	>212.0	>212.0	11/21/2022	14:20
3	N5734-02		1	13	100.00	>212.0	>212.0	11/21/2022	14:50
4	N5734-02DUP		1	14	100.00	>212.0	>212.0	11/21/2022	15:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6122972

WorkList Name : fp11-21

WorkList ID : 165203

Department : Wet-Chemistry

Date : 11-21-2022 08:34:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/01/2022	Water	N5732-01	Flash Point	Cool 4 deg C	ATCE02	K11	9DMC986SP	11/17/2022	1010B
11/25/2022	Water	N5734-02	Flash Point	Cool 4 deg C	PSEG03	K11	COMP-1	11/18/2022	1010B

Date/Time 11/21/2022 13:45
 Raw Sample Received by: RM (WC)
 Raw Sample Relinquished by: CP (SM)

Date/Time 11/21/2022 16:00
 Raw Sample Received by: CP (SM)
 Raw Sample Relinquished by: RM (WC)