


23122987

7471B

INSTRUMENT ID: CVL

Linear



A= 0.0000e+000

B= 2.0897e-004

C= -1.1903e-003

Rho= 0.9998447

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.069	0.069	335	0.000	335					
0.2	0.200	0.202	0.002	970	0.0 %	970					
2.5	2.500	2.394	-0.106	11464	0.0 %	11464					
5.0	5.000	4.978	-0.022	23829	0.0 %	23829					
7.5	7.500	7.589	0.089	36324	0.0 %	36324					
10.0	10.000	9.968	-0.032	47706	0.0 %	47706					

1
-4
0
1
0

LB122987 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 Nov 2022 10:35:33

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	335	-	0.0000	7471B	PPB	22 Nov 2022 10:44:20	S	Std
0.2 - 1	50.2	970	-	0.2000	7471B	PPB	22 Nov 2022 10:46:38	S	Std
2.5 - 1	52.5	11464	-	2.5000	7471B	PPB	22 Nov 2022 10:48:55	S	Std
5.0 - 1	55	23829	-	5.0000	7471B	PPB	22 Nov 2022 10:51:12	S	Std
7.5 - 1	57.5	36324	-	7.5000	7471B	PPB	22 Nov 2022 10:53:29	S	Std
10.0 - 1	510	47706	-	10.0000	7471B	PPB	22 Nov 2022 10:59:51	S	Std
ICV30 - 1	ICV30	18001	3.7604	-	7471B	PPB	22 Nov 2022 11:02:40	U	SMPL
ICB30 - 1	ICB30	10	0.0009	-	7471B	PPB	22 Nov 2022 11:04:55	U	SMPL
CCV64 - 1	CCV64	22169	4.6314	-	7471B	PPB	22 Nov 2022 11:07:10	U	SMPL
CCB64 - 1	CCB64	-78	-0.0175	-	7471B	PPB	22 Nov 2022 11:09:25	U	SMPL
CRA - 1	CRA	849	0.1762	-	7471B	PPB	22 Nov 2022 11:14:18	U	SMPL
HighStd - 1	HighStd	43257	9.0380	-	7471B	PPB	22 Nov 2022 11:19:09	U	SMPL
ChkStd - 1	ChkStd	31743	6.6320	-	7471B	PPB	22 Nov 2022 11:21:25	U	SMPL
PB149235BL - 1	PBS	-62	-0.0141	-	7471B	PPB	22 Nov 2022 11:23:42	U	SMPL
PB149235BS - 1	LCSS	16989	3.5489	-	7471B	PPB	22 Nov 2022 11:28:36	U	SMPL
N5718-01 - 1	ETGI-339	1971	0.4107	-	7471B	PPB	22 Nov 2022 11:32:46	U	SMPL
N5718-03 - 1	VNJ-228	1538	0.3202	-	7471B	PPB	22 Nov 2022 11:35:05	U	SMPL
N5723-01 - 1	OR-3-111822	777	0.1612	-	7471B	PPB	22 Nov 2022 11:37:22	U	SMPL
N5731-01 - 1	LEFT-SIDE	2502	0.5216	-	7471B	PPB	22 Nov 2022 11:39:37	U	SMPL
CCV65 - 1	CCV65	21778	4.5497	-	7471B	PPB	22 Nov 2022 11:41:52	U	SMPL
CCB65 - 1	CCB65	-218	-0.0467	-	7471B	PPB	22 Nov 2022 11:44:08	U	SMPL
N5731-02 - 1	RIGHT-SIDE	1744	0.3632	-	7471B	PPB	22 Nov 2022 11:46:24	U	SMPL
N5735-01 - 1	TR-04-111822	2153	0.4487	-	7471B	PPB	22 Nov 2022 11:48:39	U	SMPL
N5740-01 - 1	MH-12	958	0.1990	-	7471B	PPB	22 Nov 2022 11:50:54	U	SMPL
N5740-01DUP - 1	MH-12DUP	1013	0.2105	-	7471B	PPB	22 Nov 2022 11:55:54	U	SMPL
N5740-01MS - 1	MH-12MS	16533	3.4536	-	7471B	PPB	22 Nov 2022 11:58:10	U	SMPL
N5740-01MSD - 1	MH-12MSD	16824	3.5145	-	7471B	PPB	22 Nov 2022 12:00:27	U	SMPL
N5749-01 - 1	MH-82	478	0.0987	-	7471B	PPB	22 Nov 2022 12:02:43	U	SMPL
N5749-08 - 1	MH-83	945	0.1963	-	7471B	PPB	22 Nov 2022 12:05:00	U	SMPL
N5749-15 - 1	MH-83.5	7353	1.5353	-	7471B	PPB	22 Nov 2022 12:07:17	U	SMPL
N5749-22 - 1	MH-84	1433	0.2983	-	7471B	PPB	22 Nov 2022 12:09:32	U	SMPL
CCV66 - 1	CCV66	21705	4.5344	-	7471B	PPB	22 Nov 2022 12:11:48	U	SMPL
CCB66 - 1	CCB66	-26	-0.0066	-	7471B	PPB	22 Nov 2022 12:14:03	U	SMPL
N5751-01 - 1	STOCK-PILE	5233	1.0923	-	7471B	PPB	22 Nov 2022 12:16:19	U	SMPL
N5751-03 - 1	VNJ-223	3398	0.7089	-	7471B	PPB	22 Nov 2022 12:18:34	U	SMPL
N5753-01 - 1	72-11986	2513	0.5239	-	7471B	PPB	22 Nov 2022 12:20:50	U	SMPL
N5740-01LX5 - 1		40	0.0072	-	7471B	PPB	22 Nov 2022 12:28:57	U	SMPL
N5740-01A - 1		16393	3.4244	-	7471B	PPB	22 Nov 2022 12:31:17	U	SMPL
CCV67 - 1	CCV67	25402	5.3070	-	7471B	PPB	22 Nov 2022 12:51:24	U	SMPL
CCB67 - 1	CCB67	-248	-0.0530	-	7471B	PPB	22 Nov 2022 12:53:40	U	SMPL