

LB122999


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

44320



A= 0.0000e+000

B= 2.2488e-004 slope

C= 2.9585e-002 $y = 1.31 \bar{x} + 1.95$

Rho= 0.9999884

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%b)
0.00	0.000	0.026	0.026	-15	0.000	-15					
0.05	0.050					185	0				
0.20	0.200	0.189	-0.011	708	0.0 %	708					
2.50	2.500	2.485	-0.015	10917	0.0 %	10917					
5.00	5.000	4.982	-0.018	22022	0.0 %	22022					
7.50	7.500	7.522	0.022	33319	0.0 %	33319					
10.0	10.000	9.996	-0.004	44320	0.0 %	44320					

0.00
-16
-10
000

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-15	-		0 SFAM01.1	PPB	11/22/2022 14:39	Std
	0.2 S01	708	-		0.2 SFAM01.1	PPB	11/22/2022 14:42	Std
	2.5 S02	10917	-		2.5 SFAM01.1	PPB	11/22/2022 14:44	Std
	5 S03	22022	-		5 SFAM01.1	PPB	11/22/2022 14:46	Std
	7.5 S04	33319	-		7.5 SFAM01.1	PPB	11/22/2022 14:49	Std
	10 S05	44320	-		10 SFAM01.1	PPB	11/22/2022 14:51	Std
ICV031	ICV031	18531	4.1968 -		SFAM01.1	PPB	11/22/2022 14:54	SMPL
ICB031	ICB031	93	0.0505 -		SFAM01.1	PPB	11/22/2022 14:56	SMPL
CCV068	CCV068	22090	4.9972 -		SFAM01.1	PPB	11/22/2022 14:58	SMPL
CCB068	CCB068	-55	0.0172 -		SFAM01.1	PPB	11/22/2022 15:01	SMPL
PB149246BL	PBS246	-21	0.0249 -		SFAM01.1	PPB	11/22/2022 15:03	SMPL
N5600-01	MH0AF3	1423	0.3496 -		SFAM01.1	PPB	11/22/2022 15:05	SMPL
N5600-02	MH0AF3D	1354	0.3341 -		SFAM01.1	PPB	11/22/2022 15:07	SMPL
N5600-03	MH0AF3S	13115	2.9789 -		SFAM01.1	PPB	11/22/2022 15:10	SMPL
N5600-04	MH0AF4	12032	2.7353 -		SFAM01.1	PPB	11/22/2022 15:12	SMPL
N5600-05	MH0AF6	905	0.2331 -		SFAM01.1	PPB	11/22/2022 15:14	SMPL
N5600-06	MH0AF7	3020	0.7087 -		SFAM01.1	PPB	11/22/2022 15:16	SMPL
N5600-07	MH0AF8	2234	0.532 -		SFAM01.1	PPB	11/22/2022 15:19	SMPL
N5600-08	MH0AF9	1770	0.4276 -		SFAM01.1	PPB	11/22/2022 15:21	SMPL
N5600-09	MH0AG0	101581	22.873 -		SFAM01.1	PPB	11/22/2022 15:23	SMPL
N5600-10	MH0AG1	70675	15.9229 -		SFAM01.1	PPB	11/22/2022 15:26	SMPL
N5600-11	MH0AG2	13165	2.9901 -		SFAM01.1	PPB	11/22/2022 15:28	SMPL
N5600-12	MH0AG3	883683	HIGH -		SFAM01.1	PPB	11/22/2022 15:30	
N5600-13	MH0AG4	1414	0.3476 -		SFAM01.1	PPB	11/22/2022 15:32	SMPL
N5600-14	MH0AG5	14679	3.3306 -		SFAM01.1	PPB	11/22/2022 15:34	SMPL
N5600-15	MH0AG6	6078	1.3964 -		SFAM01.1	PPB	11/22/2022 15:36	SMPL
N5600-16	MH0AG7	34615	7.8138 -		SFAM01.1	PPB	11/22/2022 15:39	SMPL
N5600-17	MH0AG8	3645	0.8493 -		SFAM01.1	PPB	11/22/2022 15:41	SMPL
N5600-18	MH0AG9	29722	6.7134 -		SFAM01.1	PPB	11/22/2022 15:43	SMPL
N5600-19	MH0AH0	13045	2.9631 -		SFAM01.1	PPB	11/22/2022 15:46	SMPL
CCV069	CCV069	21172	4.7907 -		SFAM01.1	PPB	11/22/2022 15:48	SMPL
CCB069	CCB069	-179	-0.0107 -		SFAM01.1	PPB	11/22/2022 15:50	SMPL
N5600-20	MH0AH1	12502	2.841 -		SFAM01.1	PPB	11/22/2022 15:52	SMPL
N5600-21	MH0AH2	562	0.156 -		SFAM01.1	PPB	11/22/2022 15:55	SMPL
PB149247BL	PBS247	-79	0.0118 -		SFAM01.1	PPB	11/22/2022 15:57	SMPL
N5601-01	MH0AF5	11938	2.7142 -		SFAM01.1	PPB	11/22/2022 15:59	SMPL
N5601-02	MH0AF5D	12667	2.8781 -		SFAM01.1	PPB	11/22/2022 16:04	SMPL
N5601-03	MH0AF5S	24976	5.6462 -		SFAM01.1	PPB	11/22/2022 16:09	SMPL
N5601-04	MH0AH3	151088	34.0061 -		SFAM01.1	PPB	11/22/2022 16:11	SMPL
N5601-05	MH0AH4	5492	1.2646 -		SFAM01.1	PPB	11/22/2022 16:14	SMPL
N5601-06	MH0AH5	52707	11.8823 -		SFAM01.1	PPB	11/22/2022 16:16	SMPL
N5601-07	MH0AH6	2614	0.6174 -		SFAM01.1	PPB	11/22/2022 16:18	SMPL
N5601-08	MH0AH7	21158	4.7876 -		SFAM01.1	PPB	11/22/2022 16:21	SMPL
N5601-09	MH0AH8	17674	4.0041 -		SFAM01.1	PPB	11/22/2022 16:23	SMPL
N5601-10	MH0AH9	62754	14.1416 -		SFAM01.1	PPB	11/22/2022 16:25	SMPL
N5601-11	MH0AJ0	40097	9.0466 -		SFAM01.1	PPB	11/22/2022 16:27	SMPL

INSTRUMENT ID : CV1

N5601-12	MH0AJ1	799160	HIGH	-	SFAM01.1	PPB	11/22/2022 16:30
N5601-13	MH0AJ2	8875	2.0254	-	SFAM01.1	PPB	11/22/2022 16:31 SMPL
N5601-14	MH0AJ3	4382	1.015	-	SFAM01.1	PPB	11/22/2022 16:34 SMPL
N5601-15	MH0AJ4	1354	0.3341	-	SFAM01.1	PPB	11/22/2022 16:36 SMPL
N5601-16	MH0AJ5	-76	0.0125	-	SFAM01.1	PPB	11/22/2022 16:38 SMPL
N5601-17	MH0AJ6	-377	-0.0552	-	SFAM01.1	PPB	11/22/2022 16:41 SMPL
CCV070	CCV070	21076	4.7691	-	SFAM01.1	PPB	11/22/2022 16:43 SMPL
CCB070	CCB070	-371	-0.0538	-	SFAM01.1	PPB	11/22/2022 16:45 SMPL
N5600-09DLX5	MH0AG0	19543	4.4244	-	SFAM01.1	PPB	11/22/2022 16:47 SMPL
N5600-10DLX5	MH0AG1	13584	3.0843	-	SFAM01.1	PPB	11/22/2022 16:50 SMPL
N5600-12DLX10	MH0AG3	158211	35.6079	-	SFAM01.1	PPB	11/22/2022 16:52 SMPL
N5601-04DLX10	MH0AH3	15132	3.4325	-	SFAM01.1	PPB	11/22/2022 16:54 SMPL
N5601-06DLX2	MH0AH5	26133	5.9063	-	SFAM01.1	PPB	11/22/2022 16:57 SMPL
N5601-10DLX5	MH0AH9	12168	2.7659	-	SFAM01.1	PPB	11/22/2022 16:59 SMPL
N5601-12DLX10	MH0AJ1	124359	27.9953	-	SFAM01.1	PPB	11/22/2022 17:01 SMPL
N5600-12DLX100	MH0AG3	15588	3.535	-	SFAM01.1	PPB	11/22/2022 17:03 SMPL
N5601-12DLX50	MH0AJ1	23231	5.2537	-	SFAM01.1	PPB	11/22/2022 17:06 SMPL
CCV071	CCV071	21678	4.9045	-	SFAM01.1	PPB	11/22/2022 17:08 SMPL
CCB071	CCB071	-122	0.0021	-	SFAM01.1	PPB	11/22/2022 17:10 SMPL