


LB123008

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.2974e-004

C= 3.5637e-002

Rho= 0.9998665

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.045	0.045	42	0.000	42					
0.2	0.200	0.215	0.015	782	0.0 %	782					1
2.5	2.500	2.377	-0.123	10193	0.0 %	10193					7
5.0	5.000	5.067	0.067	21899	0.0 %	21899					-5
7.5	7.500	7.502	0.002	32498	0.0 %	32498					1
10.0	10.000	9.993	-0.007	43343	0.0 %	43343					0

LB123008 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Nov 2022 08:54:02

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	42	-	0.0000	7470A	PPB	23 Nov 2022 08:58:35	S	Std
0.2 - 1	50.2	782	-	0.2000	7470A	PPB	23 Nov 2022 09:00:53	S	Std
2.5 - 1	52.5	10193	-	2.5000	7470A	PPB	23 Nov 2022 09:03:10	S	Std
5.0 - 1	55	21899	-	5.0000	7470A	PPB	23 Nov 2022 09:05:27	S	Std
7.5 - 1	57.5	32498	-	7.5000	7470A	PPB	23 Nov 2022 09:07:43	S	Std
10.0 - 1	59.5	43343	-	10.0000	7470A	PPB	23 Nov 2022 09:12:34	S	Std
ICV32 - 1	ICV32	17017	3.9452	-7470A	PPB	23 Nov 2022 09:18:13	U	SMPL	
ICB32 - 1	ICB32	-22	0.0306	-7470A	PPB	23 Nov 2022 09:20:28	U	SMPL	
CCV72 - 1	CCV72	23028	5.3262	-7470A	PPB	23 Nov 2022 09:22:43	U	SMPL	
CCB72 - 1	CCB72	-142	0.0030	-7470A	PPB	23 Nov 2022 09:24:58	U	SMPL	
CRA - 1	CRA	706	0.1978	-7470A	PPB	23 Nov 2022 09:27:14	U	SMPL	
HighStd - 1	HighStd	43759	10.0890	-7470A	PPB	23 Nov 2022 09:29:29	U	SMPL	
ChkStd - 1	ChkStd	31086	7.1774	-7470A	PPB	23 Nov 2022 09:31:44	U	SMPL	
PB149248BL - 1	PBW	-60	0.0219	-7470A	PPB	23 Nov 2022 09:34:00	U	SMPL	
PB149248BS - 1	LCSW	18499	4.2857	-7470A	PPB	23 Nov 2022 09:36:15	U	SMPL	
N5734-01 - 1	5683	-53	0.0235	-7470A	PPB	23 Nov 2022 09:38:32	U	SMPL	
N5734-06 - 1	5617	-16	0.0320	-7470A	PPB	23 Nov 2022 09:40:48	U	SMPL	
N5734-07 - 1	5686	206	0.0830	-7470A	PPB	23 Nov 2022 09:43:06	U	SMPL	
N5740-02 - 1	MH-12	-99	0.0129	-7470A	PPB	23 Nov 2022 09:45:23	U	SMPL	
N5749-02 - 1	MH-82	-50	0.0241	-7470A	PPB	23 Nov 2022 09:47:38	U	SMPL	
CCV73 - 1	CCV73	20989	4.8577	-7470A	PPB	23 Nov 2022 09:49:54	U	SMPL	
CCB73 - 1	CCB73	-79	0.0175	-7470A	PPB	23 Nov 2022 09:52:09	U	SMPL	
N5749-09 - 1	MH-83	28	0.0421	-7470A	PPB	23 Nov 2022 09:54:25	U	SMPL	
N5749-16 - 1	MH-83.5	28	0.0421	-7470A	PPB	23 Nov 2022 09:56:40	U	SMPL	
N5749-23 - 1	MH-84	-11	0.0331	-7470A	PPB	23 Nov 2022 09:58:55	U	SMPL	
N5751-04 - 1	VNJ-223	-3	0.0349	-7470A	PPB	23 Nov 2022 10:01:11	U	SMPL	
N5754-02 - 1	PSC-124019	-8	0.0338	-7470A	PPB	23 Nov 2022 10:03:26	U	SMPL	
N5754-02DUP - 1	PSC-124019DUP	-5	0.0345	-7470A	PPB	23 Nov 2022 10:05:43	U	SMPL	
N5754-02MS - 1	PSC-124019MS	17917	4.1519	-7470A	PPB	23 Nov 2022 10:07:59	U	SMPL	
N5754-02MSD - 1	PSC-124019MSD	17782	4.1209	-7470A	PPB	23 Nov 2022 10:10:16	U	SMPL	
PB149249BL - 1	PBW	-9	0.0336	-7470A	PPB	23 Nov 2022 10:12:33	U	SMPL	
PB149249BS - 1	LCSW	19620	4.5432	-7470A	PPB	23 Nov 2022 10:14:49	U	SMPL	
CCV74 - 1	CCV74	23170	5.3588	-7470A	PPB	23 Nov 2022 10:17:05	U	SMPL	
CCB74 - 1	CCB74	-26	0.0297	-7470A	PPB	23 Nov 2022 10:19:20	U	SMPL	
N5732-01 - 1	9DMC986SP	168	0.0742	-7470A	PPB	23 Nov 2022 10:21:36	U	SMPL	
N5734-02 - 1	COMP-1	47	0.0464	-7470A	PPB	23 Nov 2022 10:23:52	U	SMPL	
N5734-02DUP - 1	COMP-1DUP	2	0.0361	-7470A	PPB	23 Nov 2022 10:26:07	U	SMPL	
N5734-02MS - 1	COMP-1MS	16720	3.8768	-7470A	PPB	23 Nov 2022 10:28:23	U	SMPL	
N5734-02MSD - 1	COMP-1MSD	16754	3.8848	-7470A	PPB	23 Nov 2022 10:30:39	U	SMPL	
PB149250BL - 1	PBW	-12	0.0329	-7470A	PPB	23 Nov 2022 10:32:55	U	SMPL	
PB149250BS - 1	LCSW	18404	4.2638	-7470A	PPB	23 Nov 2022 10:35:11	U	SMPL	
N5675-02 - 1	AUD-1914-TOTE	824	0.2249	-7470A	PPB	23 Nov 2022 10:42:26	U	SMPL	
N5675-02DUP - 1	AUD-1914-TOTEDUP	738	0.2052	-7470A	PPB	23 Nov 2022 10:44:42	U	SMPL	
N5675-02MS - 1	AUD-1914-TOTEMS	17857	4.1382	-7470A	PPB	23 Nov 2022 10:46:58	U	SMPL	
CCV75 - 1	CCV75	22178	5.1309	-7470A	PPB	23 Nov 2022 10:49:14	U	SMPL	
CCB75 - 1	CCB75	-5	0.0345	-7470A	PPB	23 Nov 2022 10:51:29	U	SMPL	
N5675-02MSD - 1	AUD-1914-TOTEMSD	17618	4.0833	-7470A	PPB	23 Nov 2022 10:53:45	U	SMPL	
N5675-04 - 1	AUD-1742-TOTE	7588	1.7789	-7470A	PPB	23 Nov 2022 10:56:01	U	SMPL	
N5675-06 - 1	AUD-1912	32	0.0430	-7470A	PPB	23 Nov 2022 10:58:16	U	SMPL	
N5709-01 - 1	COMP-SAMPLE-TANK-12-31-32-33	109	0.0607	-7470A	PPB	23 Nov 2022 11:00:32	U	SMPL	
N5752-25 - 1	TP-1	5754	1.3576	-7470A	PPB	23 Nov 2022 11:02:48	U	SMPL	
N5752-26 - 1	TP-1	-61	0.0216	-7470A	PPB	23 Nov 2022 11:05:03	U	SMPL	
PB149225TB - 1	PB149225TB	-5	0.0345	-7470A	PPB	23 Nov 2022 11:07:20	U	SMPL	
PB149231TB - 1	PB149231TB	15	0.0391	-7470A	PPB	23 Nov 2022 11:09:38	U	SMPL	
CCV76 - 1	CCV76	22253	5.1481	-7470A	PPB	23 Nov 2022 11:11:55	U	SMPL	
CCB76 - 1	CCB76	-22	0.0306	-7470A	PPB	23 Nov 2022 11:14:11	U	SMPL	
N5754-02LX5 - 1		-14	0.0324	-7470A	PPB	23 Nov 2022 11:16:28	U	SMPL	
N5754-02A - 1		18238	4.2257	-7470A	PPB	23 Nov 2022 11:18:44	U	SMPL	
N5734-02LX5 - 1		-66	0.0205	-7470A	PPB	23 Nov 2022 11:20:59	U	SMPL	
N5734-02A - 1		16898	3.9178	-7470A	PPB	23 Nov 2022 11:23:15	U	SMPL	
N5675-02LX5 - 1		128	0.0650	-7470A	PPB	23 Nov 2022 11:25:30	U	SMPL	
N5675-02A - 1		17154	3.9767	-7470A	PPB	23 Nov 2022 11:27:46	U	SMPL	
CCV77 - 1	CCV77	21722	5.0261	-7470A	PPB	23 Nov 2022 11:31:48	U	SMPL	
CCB77 - 1	CCB77	11	0.0382	-7470A	PPB	23 Nov 2022 11:34:04	U	SMPL	