

LB123013


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

53233



A= 0.0000e+000

B= 1.8813e-004

C= 1.1318e-002

Rho= 0.9999565

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.041	0.041	158	0.000	158					7
0.2	0.200	0.214	0.014	1076	0.0 %	1076					-3
2.5	2.500	2.434	-0.066	12876	0.0 %	12876					0
5.0	5.000	4.998	-0.002	26508	0.0 %	26508					0
7.5	7.500	7.488	-0.012	39741	0.0 %	39741					0
10.0	10.000	10.026	0.026	53233	0.0 %	53233					0

LB123013 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 23 Nov 2022 13:28:53

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	158	-	0.0000	7471B	PPB	23 Nov 2022 13:33:38	S	Std
0.2 - 1	50.2	1076	-	0.2000	7471B	PPB	23 Nov 2022 13:35:55	S	Std
2.5 - 1	52.5	12876	-	2.5000	7471B	PPB	23 Nov 2022 13:38:12	S	Std
5.0 - 1	55	26508	-	5.0000	7471B	PPB	23 Nov 2022 13:40:29	S	Std
7.5 - 1	57.5	39741	-	7.5000	7471B	PPB	23 Nov 2022 13:42:45	S	Std
10.0 - 1	510	53233	-	10.0000	7471B	PPB	23 Nov 2022 13:47:41	S	Std
ICV33 - 1	ICV33	21157	3.9915	-	7471B	PPB	23 Nov 2022 13:50:28	U	SMPL
ICB33 - 1	ICB33	205	0.0499	-	7471B	PPB	23 Nov 2022 13:52:43	U	SMPL
CCV78 - 1	CCV78	26283	4.9558	-	7471B	PPB	23 Nov 2022 13:54:58	U	SMPL
CCB78 - 1	CCB78	76	0.0256	-	7471B	PPB	23 Nov 2022 13:57:13	U	SMPL
CRA - 1	CRA	1081	0.2147	-	7471B	PPB	23 Nov 2022 13:59:28	U	SMPL
HighStd - 1	HighStd	53051	9.9916	-	7471B	PPB	23 Nov 2022 14:01:44	U	SMPL
ChkStd - 1	ChkStd	37110	6.9927	-	7471B	PPB	23 Nov 2022 14:03:59	U	SMPL
PB149278BL - 1	PBS	-37	0.0044	-	7471B	PPB	23 Nov 2022 14:06:16	U	SMPL
PB149278BS - 1	LCSS	18954	3.5771	-	7471B	PPB	23 Nov 2022 14:08:32	U	SMPL
N5317-11 - 1	PT-MET-SOIL	643430	HIGH	-	7471B	PPB	23 Nov 2022 14:10:48	X	
N5745-01 - 1	PAT-01-112122	1551	0.3031	-	7471B	PPB	23 Nov 2022 14:12:38	U	SMPL
N5745-03 - 1	PAT-04-112122	1752	0.3409	-	7471B	PPB	23 Nov 2022 14:14:58	U	SMPL
N5754-01 - 1	PSC-124019	2469	0.4758	-	7471B	PPB	23 Nov 2022 14:17:15	U	SMPL
N5758-01 - 1	MH-13	1164	0.2303	-	7471B	PPB	23 Nov 2022 14:19:30	U	SMPL
CCV79 - 1	CCV79	26304	4.9598	-	7471B	PPB	23 Nov 2022 14:21:46	U	SMPL
CCB79 - 1	CCB79	-81	-0.0039	-	7471B	PPB	23 Nov 2022 14:24:02	U	SMPL
N5762-05 - 1	TP-1	725	0.1477	-	7471B	PPB	23 Nov 2022 14:26:18	U	SMPL
N5762-08 - 1	TP-2	592	0.1227	-	7471B	PPB	23 Nov 2022 14:28:33	U	SMPL
N5765-01 - 1	TP-L	1017	0.2026	-	7471B	PPB	23 Nov 2022 14:30:49	U	SMPL
N5765-04 - 1	MH-14	127	0.0352	-	7471B	PPB	23 Nov 2022 14:33:04	U	SMPL
N5766-01 - 1	CR-1	1825	0.3546	-	7471B	PPB	23 Nov 2022 14:35:20	U	SMPL
N5766-02 - 1	CR-2	651	0.1338	-	7471B	PPB	23 Nov 2022 14:37:36	U	SMPL
N5766-03 - 1	CR-3	1562	0.3052	-	7471B	PPB	23 Nov 2022 14:39:52	U	SMPL
N5766-04 - 1	CR-4	356	0.0783	-	7471B	PPB	23 Nov 2022 14:42:09	U	SMPL
N5766-05 - 1	CR-5	726	0.1479	-	7471B	PPB	23 Nov 2022 14:44:26	U	SMPL
N5766-05DUP - 1	CR-5DUP	728	0.1483	-	7471B	PPB	23 Nov 2022 14:46:42	U	SMPL
CCV80 - 1	CCV80	26334	4.9654	-	7471B	PPB	23 Nov 2022 14:48:57	U	SMPL
CCB80 - 1	CCB80	37	0.0183	-	7471B	PPB	23 Nov 2022 14:51:13	U	SMPL
N5766-05MS - 1	CR-5MS	21745	4.1021	-	7471B	PPB	23 Nov 2022 14:53:29	U	SMPL
N5766-05MSD - 1	CR-5MSD	20634	3.8931	-	7471B	PPB	23 Nov 2022 14:55:44	U	SMPL
N5766-05LX5 - 1		69	0.0243	-	7471B	PPB	23 Nov 2022 14:58:00	U	SMPL
N5766-05A - 1		21744	4.1019	-	7471B	PPB	23 Nov 2022 15:00:16	U	SMPL
N5317-11DL2X10 - 1	PT-MET-SOIL	101425	19.0920	-	7471B	PPB	23 Nov 2022 15:09:08	U	SMPL
N5317-11DL2X50 - 1	PT-MET-SOIL	20456	3.8596	-	7471B	PPB	23 Nov 2022 15:13:42	U	SMPL
CCV81 - 1	CCV81	26049	4.9118	-	7471B	PPB	23 Nov 2022 15:15:58	U	SMPL
CCB81 - 1	CCB81	-153	-0.0175	-	7471B	PPB	23 Nov 2022 15:18:14	U	SMPL