

LB123025


7470A

INSTRUMENT ID: CV2

Linear

 $\mu$  Abs.:

46252



A= 0.0000e+000

B= 2.1434e-004

C= 3.3143e-002

Rho= 0.9998812

Accept=Accepted

| Std ID | Conc.  | Calc. | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %1) |
|--------|--------|-------|--------|-------|------------|-------|-------|-------|-------|-------|-----|
| 0.0    | 0.000  | 0.029 | 0.029  | -21   | 0.000      | -21   |       |       |       |       | 7   |
| 0.2    | 0.200  | 0.188 | -0.012 | 722   | 0.0 %      | 722   |       |       |       |       | -6  |
| 2.5    | 2.500  | 2.413 | -0.087 | 11105 | 0.0 %      | 11105 |       |       |       |       | -3  |
| 5.0    | 5.000  | 5.077 | 0.077  | 23533 | 0.0 %      | 23533 |       |       |       |       | 2   |
| 7.5    | 7.500  | 7.546 | 0.046  | 35053 | 0.0 %      | 35053 |       |       |       |       | 1   |
| 10.0   | 10.000 | 9.947 | -0.053 | 46252 | 0.0 %      | 46252 |       |       |       |       | -1  |

## LB123025 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 25 Nov 2022 10:50:31

| Sample ID       | Extended ID | $\mu$ Abs. | Conc.  | Std Conc | Method | Units | Date                 | Type | Type |
|-----------------|-------------|------------|--------|----------|--------|-------|----------------------|------|------|
| 0.0 - 1         | S0          | -21        | -      | 0.0000   | 7470A  | PPB   | 25 Nov 2022 10:58:03 | S    | Std  |
| 0.2 - 1         | S0-2        | 722        | -      | 0.2000   | 7470A  | PPB   | 25 Nov 2022 11:00:20 | S    | Std  |
| 2.5 - 1         | S2-5        | 11105      | -      | 2.5000   | 7470A  | PPB   | 25 Nov 2022 11:02:38 | S    | Std  |
| 5.0 - 1         | S5          | 23533      | -      | 5.0000   | 7470A  | PPB   | 25 Nov 2022 11:04:55 | S    | Std  |
| 7.5 - 1         | S7-5        | 35053      | -      | 7.5000   | 7470A  | PPB   | 25 Nov 2022 11:07:12 | S    | Std  |
| 10.0 - 1        | S10         | 46252      | -      | 10.0000  | 7470A  | PPB   | 25 Nov 2022 11:09:28 | S    | Std  |
| ICV34 - 1       | ICV34       | 17979      | 3.8867 | -        | 7470A  | PPB   | 25 Nov 2022 11:12:12 | U    | SMPL |
| ICB34 - 1       | ICB34       | 161        | 0.0677 | -        | 7470A  | PPB   | 25 Nov 2022 11:14:26 | U    | SMPL |
| CCV82 - 1       | CCV82       | 21446      | 4.6298 | -        | 7470A  | PPB   | 25 Nov 2022 11:19:15 | U    | SMPL |
| CCB82 - 1       | CCB82       | -77        | 0.0166 | -        | 7470A  | PPB   | 25 Nov 2022 11:21:30 | U    | SMPL |
| CRA - 1         | CRA         | 802        | 0.2050 | -        | 7470A  | PPB   | 25 Nov 2022 11:23:45 | U    | SMPL |
| HighStd - 1     | HighStd     | 45908      | 9.8729 | -        | 7470A  | PPB   | 25 Nov 2022 11:26:00 | U    | SMPL |
| ChkStd - 1      | ChkStd      | 29517      | 6.3597 | -        | 7470A  | PPB   | 25 Nov 2022 11:28:15 | U    | SMPL |
| PB149279BL - 1  | PBW         | 34         | 0.0404 | -        | 7470A  | PPB   | 25 Nov 2022 11:30:31 | U    | SMPL |
| PB149279BS - 1  | LCSW        | 16722      | 3.6173 | -        | 7470A  | PPB   | 25 Nov 2022 11:32:47 | U    | SMPL |
| N5758-02 - 1    | MH-13       | 47         | 0.0432 | -        | 7470A  | PPB   | 25 Nov 2022 11:35:03 | U    | SMPL |
| N5762-06 - 1    | TP-1        | 33         | 0.0402 | -        | 7470A  | PPB   | 25 Nov 2022 11:37:19 | U    | SMPL |
| N5762-09 - 1    | TP-2        | 17         | 0.0368 | -        | 7470A  | PPB   | 25 Nov 2022 11:39:36 | U    | SMPL |
| N5764-01 - 1    | PCD-304     | 22         | 0.0379 | -        | 7470A  | PPB   | 25 Nov 2022 11:41:54 | U    | SMPL |
| N5765-02 - 1    | TP-L        | 96         | 0.0537 | -        | 7470A  | PPB   | 25 Nov 2022 11:44:11 | U    | SMPL |
| CCV83 - 1       | CCV83       | 22689      | 4.8962 | -        | 7470A  | PPB   | 25 Nov 2022 11:46:26 | U    | SMPL |
| CCB83 - 1       | CCB83       | -38        | 0.0250 | -        | 7470A  | PPB   | 25 Nov 2022 11:48:41 | U    | SMPL |
| N5765-05 - 1    | MH-14       | 157        | 0.0668 | -        | 7470A  | PPB   | 25 Nov 2022 11:50:56 | U    | SMPL |
| N5765-05DUP - 1 | MH-14DUP    | 139        | 0.0629 | -        | 7470A  | PPB   | 25 Nov 2022 11:53:11 | U    | SMPL |
| N5765-05MS - 1  | MH-14MS     | 15696      | 3.3974 | -        | 7470A  | PPB   | 25 Nov 2022 11:55:27 | U    | SMPL |
| N5765-05MSD - 1 | MH-14MSD    | 17873      | 3.8640 | -        | 7470A  | PPB   | 25 Nov 2022 11:57:43 | U    | SMPL |
| N5765-05LX5 - 1 |             | 59         | 0.0458 | -        | 7470A  | PPB   | 25 Nov 2022 12:02:49 | U    | SMPL |
| N5765-05A - 1   |             | 18492      | 3.9967 | -        | 7470A  | PPB   | 25 Nov 2022 12:05:05 | U    | SMPL |
| PB149244TB - 1  | PB149244TB  | -17        | 0.0295 | -        | 7470A  | PPB   | 25 Nov 2022 12:07:22 | U    | SMPL |
| CCV84 - 1       | CCV84       | 22562      | 4.8690 | -        | 7470A  | PPB   | 25 Nov 2022 12:09:39 | U    | SMPL |
| CCB84 - 1       | CCB84       | 16         | 0.0366 | -        | 7470A  | PPB   | 25 Nov 2022 12:11:56 | U    | SMPL |