

LB123026


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

53952



A= 0.0000e+000

B= 1.8747e-004

C= -1.2884e-002

Rho= 0.9997641

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/d1)
0.0	0.000	-0.012	-0.012	6	0.000	6					1
0.2	0.200	0.213	0.013	1203	0.0 %	1203					6
2.5	2.500	2.563	0.063	13740	0.0 %	13740					3
5.0	5.000	4.989	-0.011	26683	0.0 %	26683					0
7.5	7.500	7.345	-0.155	39251	0.0 %	39251					-2
10.0	10.000	10.101	0.101	53952	0.0 %	53952					1

LB123026 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 25 Nov 2022 13:57:45

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	6	-	0.00007471B	PPB		25 Nov 2022 13:59:17	S	Std
0.2 - 1	50.2	1203	-	0.20007471B	PPB		25 Nov 2022 14:01:34	S	Std
2.5 - 1	52.5	13740	-	2.50007471B	PPB		25 Nov 2022 14:03:51	S	Std
5.0 - 1	55	26683	-	5.00007471B	PPB		25 Nov 2022 14:06:08	S	Std
7.5 - 1	57.5	39251	-	7.50007471B	PPB		25 Nov 2022 14:08:24	S	Std
10.0 - 1	510	53952	-	10.00007471B	PPB		25 Nov 2022 14:10:40	S	Std
ICV35 - 1	ICV35	21599	4.0362	-7471B	PPB		25 Nov 2022 14:13:30	U	SMPL
ICB35 - 1	ICB35	41	-0.0052	-7471B	PPB		25 Nov 2022 14:15:45	U	SMPL
CCV85 - 1	CCV85	26008	4.8628	-7471B	PPB		25 Nov 2022 14:18:01	U	SMPL
CCB85 - 1	CCB85	-54	-0.0230	-7471B	PPB		25 Nov 2022 14:20:16	U	SMPL
CRA - 1	CRA	1168	0.2061	-7471B	PPB		25 Nov 2022 14:22:31	U	SMPL
HighStd - 1	HighStd	53918	10.0950	-7471B	PPB		25 Nov 2022 14:24:46	U	SMPL
ChkStd - 1	ChkStd	36872	6.8995	-7471B	PPB		25 Nov 2022 14:27:01	U	SMPL
PB149291BL - 1	PBS	-104	-0.0324	-7471B	PPB		25 Nov 2022 14:29:19	U	SMPL
PB149291BS - 1	LCSS	19796	3.6982	-7471B	PPB		25 Nov 2022 14:31:34	U	SMPL
N5752-01 - 1	B-7S	19589	3.6594	-7471B	PPB		25 Nov 2022 14:33:51	U	SMPL
N5752-02 - 1	B-7D	11142	2.0759	-7471B	PPB		25 Nov 2022 14:36:07	U	SMPL
N5752-03 - 1	B-8S	40437	7.5678	-7471B	PPB		25 Nov 2022 14:38:24	U	SMPL
N5752-04 - 1	B-8D	20524	3.8347	-7471B	PPB		25 Nov 2022 14:40:41	U	SMPL
N5752-05 - 1	B-9S	5459	1.0105	-7471B	PPB		25 Nov 2022 14:42:56	U	SMPL
CCV86 - 1	CCV86	25994	4.8602	-7471B	PPB		25 Nov 2022 14:45:12	U	SMPL
CCB86 - 1	CCB86	-70	-0.0260	-7471B	PPB		25 Nov 2022 14:47:27	U	SMPL
N5752-06 - 1	B-9D	14660	2.7354	-7471B	PPB		25 Nov 2022 14:49:42	U	SMPL
N5752-07 - 1	B-10S	1008484	HIGH	-7471B	PPB		25 Nov 2022 14:51:57	X	
N5752-08 - 1	B-10D	43449	8.1324	-7471B	PPB		25 Nov 2022 14:53:45	U	SMPL
N5752-09 - 1	B-11S	48846	9.1442	-7471B	PPB		25 Nov 2022 14:56:13	U	SMPL
N5752-10 - 1	B-11D	13092	2.4415	-7471B	PPB		25 Nov 2022 14:58:29	U	SMPL
N5752-11 - 1	B-12S	60767	11.3790	-7471B	PPB		25 Nov 2022 15:00:45	U	SMPL
N5752-12 - 1	B-12D	2344	0.4265	-7471B	PPB		25 Nov 2022 15:03:02	U	SMPL
N5752-13 - 1	B-13S	451255	HIGH	-7471B	PPB		25 Nov 2022 15:05:20	X	
N5752-14 - 1	B-13D	3387	0.6221	-7471B	PPB		25 Nov 2022 15:07:11	U	SMPL
N5752-15 - 1	B-14S	436700	HIGH	-7471B	PPB		25 Nov 2022 15:09:29	X	
CCV87 - 1	CCV87	25845	4.8322	-7471B	PPB		25 Nov 2022 15:11:18	U	SMPL
CCB87 - 1	CCB87	-47	-0.0217	-7471B	PPB		25 Nov 2022 15:13:35	U	SMPL
N5752-16 - 1	B-14D	13493	2.5166	-7471B	PPB		25 Nov 2022 15:15:51	U	SMPL
N5752-17 - 1	B-15S	684984	HIGH	-7471B	PPB		25 Nov 2022 15:18:07	X	
N5752-18 - 1	B-15D	72792	13.6333	-7471B	PPB		25 Nov 2022 15:19:55	U	SMPL
N5752-20 - 1	B-16D	17193	3.2103	-7471B	PPB		25 Nov 2022 15:22:14	U	SMPL
N5752-20DUP - 1	B-16DDUP	19029	3.5545	-7471B	PPB		25 Nov 2022 15:27:20	U	SMPL
N5752-20MS - 1	B-16DMS	36312	6.7945	-7471B	PPB		25 Nov 2022 15:32:31	U	SMPL
N5752-20MSD - 1	B-16DMSD	35925	6.7219	-7471B	PPB		25 Nov 2022 15:34:48	U	SMPL
PB149292BL - 1	PBS	-252	-0.0601	-7471B	PPB		25 Nov 2022 15:37:05	U	SMPL
PB149292BS - 1	LCSS	19240	3.5940	-7471B	PPB		25 Nov 2022 15:39:23	U	SMPL
N5769-01 - 1	COMP-218	1174	0.2072	-7471B	PPB		25 Nov 2022 15:41:39	U	SMPL
CCV88 - 1	CCV88	24572	4.5936	-7471B	PPB		25 Nov 2022 15:43:54	U	SMPL
CCB88 - 1	CCB88	-179	-0.0464	-7471B	PPB		25 Nov 2022 15:46:10	U	SMPL
N5772-01 - 1	LOCATION-1	81673	15.2982	-7471B	PPB		25 Nov 2022 15:48:25	U	SMPL
N5772-02 - 1	LOCATION-2	14338	2.6750	-7471B	PPB		25 Nov 2022 15:50:41	U	SMPL
N5772-03 - 1	LOCATION-3	40390	7.5590	-7471B	PPB		25 Nov 2022 15:52:58	U	SMPL
N5772-04 - 1	LOCATION-4	19702	3.6806	-7471B	PPB		25 Nov 2022 15:55:14	U	SMPL
N5779-01 - 1	CL-01-112322	5592	1.0354	-7471B	PPB		25 Nov 2022 15:57:30	U	SMPL
N5780-01 - 1	TP-M	1136	0.2001	-7471B	PPB		25 Nov 2022 15:59:46	U	SMPL
N5780-01DUP - 1	TP-MDUP	1117	0.1965	-7471B	PPB		25 Nov 2022 16:02:03	U	SMPL
N5780-01MS - 1	TP-MMS	19243	3.5946	-7471B	PPB		25 Nov 2022 16:04:20	U	SMPL
N5780-01MSD - 1	TP-MMSD	21130	3.9483	-7471B	PPB		25 Nov 2022 16:06:37	U	SMPL
CCV89 - 1	CCV89	25458	4.7597	-7471B	PPB		25 Nov 2022 16:08:53	U	SMPL
CCB89 - 1	CCB89	-107	-0.0329	-7471B	PPB		25 Nov 2022 16:11:09	U	SMPL
N5752-20LX5 - 1		3007	0.5508	-7471B	PPB		25 Nov 2022 16:13:24	U	SMPL
N5752-20A - 1		37540	7.0247	-7471B	PPB		25 Nov 2022 16:15:40	U	SMPL
N5780-01LX5 - 1		20	-0.0091	-7471B	PPB		25 Nov 2022 16:17:56	U	SMPL
N5780-01A - 1		20386	3.8088	-7471B	PPB		25 Nov 2022 16:20:12	U	SMPL
N5752-07DLX10 - 1		180160	33.7614	-7471B	PPB		25 Nov 2022 16:22:27	U	SMPL
N5752-11DLX2 - 1		29053	5.4336	-7471B	PPB		25 Nov 2022 16:24:43	U	SMPL
N5752-13DLX10 - 1		62717	11.7446	-7471B	PPB		25 Nov 2022 16:27:02	U	SMPL
N5752-15DLX10 - 1		58084	10.8760	-7471B	PPB		25 Nov 2022 16:29:18	U	SMPL
N5752-17DLX10 - 1		101056	18.9319	-7471B	PPB		25 Nov 2022 16:31:36	U	SMPL
N5752-18DLX5 - 1		14057	2.6224	-7471B	PPB		25 Nov 2022 16:33:53	U	SMPL
CCV90 - 1	CCV90	25011	4.6759	-7471B	PPB		25 Nov 2022 16:36:11	U	SMPL
CCB90 - 1	CCB90	-42	-0.0208	-7471B	PPB		25 Nov 2022 16:38:28	U	SMPL
N5772-01DLX5 - 1		15814	2.9517	-7471B	PPB		25 Nov 2022 16:40:43	U	SMPL
N5752-07DL2X100 - 1		17474	3.2629	-7471B	PPB		25 Nov 2022 16:43:00	U	SMPL
N5752-13DL2X25 - 1		24560	4.5913	-7471B	PPB		25 Nov 2022 16:45:16	U	SMPL
N5752-15DL2X25 - 1		23230	4.3420	-7471B	PPB		25 Nov 2022 16:47:32	U	SMPL
N5752-17DL2X50 - 1		19835	3.7056	-7471B	PPB		25 Nov 2022 16:49:48	U	SMPL
CCV91 - 1	CCV91	25086	4.6899	-7471B	PPB		25 Nov 2022 16:52:05	U	SMPL
CCB91 - 1	CCB91	-113	-0.0341	-7471B	PPB		25 Nov 2022 16:54:21	U	SMPL