

LB123040


7470A

INSTRUMENT ID: CV4

Linear

 μ Abs.:

47548



A= 0.0000e+000

B= 2.0799e-004

C= -1.3675e-002

Rho= 0.9996975

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.016	0.016	144	0.000	144					-
0.2	0.200	0.187	-0.013	967	0.0 %	967					-7
2.5	2.500	2.432	-0.068	11757	0.0 %	11757					-3
5.0	5.000	5.020	0.020	24202	0.0 %	24202					0
7.5	7.500	7.669	0.169	36937	0.0 %	36937					2
10.0	10.000	9.876	-0.124	47548	0.0 %	47548					-1

LB123040 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Nov 2022 12:50:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	144	-	0.0000	7470A	PPB	28 Nov 2022 13:30:03	S	Std
0.2 - 1	50.2	967	-	0.2000	7470A	PPB	28 Nov 2022 13:32:21	S	Std
2.5 - 1	52.5	11757	-	2.5000	7470A	PPB	28 Nov 2022 13:34:38	S	Std
5.0 - 1	55	24202	-	5.0000	7470A	PPB	28 Nov 2022 13:36:55	S	Std
7.5 - 1	57.5	36937	-	7.5000	7470A	PPB	28 Nov 2022 13:39:11	S	Std
10.0 - 1	510	47548	-	10.0000	7470A	PPB	28 Nov 2022 13:41:27	S	Std
ICV36 - 1	ICV36	18038	3.7380	-	7470A	PPB	28 Nov 2022 13:44:13	U	SMPL
ICB36 - 1	ICB36	47	-0.0039	-	7470A	PPB	28 Nov 2022 13:49:00	U	SMPL
CCV90 - 1	CCV90	23913	4.9600	-	7470A	PPB	28 Nov 2022 13:51:14	U	SMPL
CCB90 - 1	CCB90	84	0.0038	-	7470A	PPB	28 Nov 2022 13:53:30	U	SMPL
CRA - 1	CRA	911	0.1758	-	7470A	PPB	28 Nov 2022 13:55:45	U	SMPL
HighStd - 1	HighStd	47110	9.7847	-	7470A	PPB	28 Nov 2022 13:58:01	U	SMPL
ChkStd - 1	ChkStd	33694	6.9943	-	7470A	PPB	28 Nov 2022 14:02:48	U	SMPL
PB149175BL - 1	PBW	14	-0.0108	-	7470A	PPB	28 Nov 2022 14:05:03	U	SMPL
PB149175BS - 1	LCSW	20711	4.2940	-	7470A	PPB	28 Nov 2022 14:07:19	U	SMPL
N5678-02 - 1	250FB-BP-BACKFILL-01	469	0.0839	-	7470A	PPB	28 Nov 2022 14:09:35	U	SMPL
N5678-04 - 1	250FB-BP-BACKFILL-02	-2	-0.0141	-	7470A	PPB	28 Nov 2022 14:11:52	U	SMPL
N5678-06 - 1	250FB-BP-BACKFILL-03	166	0.0209	-	7470A	PPB	28 Nov 2022 14:14:09	U	SMPL
N5678-08 - 1	250FB-BP-BACKFILL-04	59	-0.0014	-	7470A	PPB	28 Nov 2022 14:16:27	U	SMPL
N5678-10 - 1	250FB-BP-BACKFILL-05	54	-0.0024	-	7470A	PPB	28 Nov 2022 14:18:42	U	SMPL
CCV91 - 1	CCV91	23586	4.8919	-	7470A	PPB	28 Nov 2022 14:20:58	U	SMPL
CCB91 - 1	CCB91	157	0.0190	-	7470A	PPB	28 Nov 2022 14:23:13	U	SMPL
N5678-12 - 1	250FB-BP-BACKFILL-06	241	0.0364	-	7470A	PPB	28 Nov 2022 14:25:29	U	SMPL
N5678-14 - 1	250FB-BP-BACKFILL-07	179	0.0236	-	7470A	PPB	28 Nov 2022 14:27:44	U	SMPL
N5678-16 - 1	250FB-BP-BACKFILL-08	94	0.0059	-	7470A	PPB	28 Nov 2022 14:29:59	U	SMPL
N5678-18 - 1	250FB-BP-BACKFILL-09	84	0.0038	-	7470A	PPB	28 Nov 2022 14:32:15	U	SMPL
N5678-20 - 1	250FB-BP-BACKFILL-10	43	-0.0047	-	7470A	PPB	28 Nov 2022 14:34:30	U	SMPL
N5678-22 - 1	250FB-BP-BACKFILL-11	30	-0.0074	-	7470A	PPB	28 Nov 2022 14:36:47	U	SMPL
N5678-22DUP - 1	250FB-BP-BACKFILL-11DUP	33	-0.0068	-	7470A	PPB	28 Nov 2022 14:39:03	U	SMPL
N5678-22MS - 1	250FB-BP-BACKFILL-11MS	17440	3.6136	-	7470A	PPB	28 Nov 2022 14:41:20	U	SMPL
N5678-22MSD - 1	250FB-BP-BACKFILL-11MSD	18161	3.7636	-	7470A	PPB	28 Nov 2022 14:43:37	U	SMPL
CCV92 - 1	CCV92	24004	4.9789	-	7470A	PPB	28 Nov 2022 14:45:53	U	SMPL
CCB92 - 1	CCB92	186	0.0250	-	7470A	PPB	28 Nov 2022 14:48:08	U	SMPL
N5678-22LX5 - 1		66	0.0001	-	7470A	PPB	28 Nov 2022 14:50:24	U	SMPL
N5678-22A - 1		19136	3.9664	-	7470A	PPB	28 Nov 2022 14:52:39	U	SMPL
PB149175 - 1	PB149145TB	78	0.0025	-	7470A	PPB	28 Nov 2022 14:54:57	U	SMPL
CCV93 - 1	CCV93	23202	4.8121	-	7470A	PPB	28 Nov 2022 14:57:13	U	SMPL
CCB93 - 1	CCB93	68	0.0005	-	7470A	PPB	28 Nov 2022 14:59:28	U	SMPL