

L8123044


245.1

INSTRUMENT ID: EV1

Linear

μ Abs.:

46147



A= 0.0000e+000

B= 2.1792e-004

C= -1.4019e-002

Rho= 0.9998966

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Med
0.0	0.000	-0.004	-0.004	45	0.000	45					5
0.05	0.050	0.048	-0.002	283	0.0 %	283	0				-4
0.2	0.200	0.179	-0.021	885	0.0 %	885					-11
2.5	2.500	2.534	0.034	11693	0.0 %	11693					1
5.0	5.000	5.063	0.063	23300	0.0 %	23300					1
7.5	7.500	7.388	-0.112	33968	0.0 %	33968					-1
10.0	10.000	10.042	0.042	46147	0.0 %	46147					0

LB123044 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 28 Nov 2022 15:53:32

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	45	-	0.0000	245.1	PPB	28 Nov 2022 16:04:02	S	Std
0.05 - 1	50.05	283	-	0.0500	245.1	PPB	28 Nov 2022 16:06:19	S	Std
0.2 - 1	50.2	885	-	0.2000	245.1	PPB	28 Nov 2022 16:08:36	S	Std
2.5 - 1	52.5	11693	-	2.5000	245.1	PPB	28 Nov 2022 16:10:53	S	Std
5.0 - 1	55	23300	-	5.0000	245.1	PPB	28 Nov 2022 16:13:10	S	Std
7.5 - 1	57.5	33968	-	7.5000	245.1	PPB	28 Nov 2022 16:15:25	S	Std
10.0 - 1	510	46147	-	10.0000	245.1	PPB	28 Nov 2022 16:20:16	S	Std
ICV37 - 1	ICV37	18066	3.9228	-245.1		PPB	28 Nov 2022 16:28:24	U	SMPL
ICB37 - 1	ICB37	-8	-0.0158	-245.1		PPB	28 Nov 2022 16:30:39	U	SMPL
CCV93 - 1	CCV93	23203	5.0423	-245.1		PPB	28 Nov 2022 16:32:55	U	SMPL
CCB93 - 1	CCB93	-44	-0.0236	-245.1		PPB	28 Nov 2022 16:35:11	U	SMPL
CRA - 1	CRA	283	0.0477	-245.1		PPB	28 Nov 2022 16:37:27	U	SMPL
HighStd - 1	HighStd	46717	10.1663	-245.1		PPB	28 Nov 2022 16:39:42	U	SMPL
ChkStd - 1	ChkStd	31125	6.7686	-245.1		PPB	28 Nov 2022 16:41:57	U	SMPL
PB149321BL - 1	PBW	74	0.0021	-245.1		PPB	28 Nov 2022 16:46:53	U	SMPL
PB149321BS - 1	LCSW	17585	3.8180	-245.1		PPB	28 Nov 2022 16:49:09	U	SMPL
N4955-10 - 1	LOD-MDL-WATER-04-QT4-2022	283	0.0477	-245.1		PPB	28 Nov 2022 16:53:59	U	SMPL
N4955-11 - 1	LOQ-WATER-05-QT4-2022	867	0.1749	-245.1		PPB	28 Nov 2022 16:56:16	U	SMPL
N5655-01 - 1	SEMI-ANNUAL-CONDENSATE	3377	0.7219	-245.1		PPB	28 Nov 2022 16:58:33	U	SMPL
N5655-01DUP - 1	SEMI-ANNUAL-CONDENSATEDUP	3297	0.7044	-245.1		PPB	28 Nov 2022 17:00:50	U	SMPL
N5655-01MS - 1	SEMI-ANNUAL-CONDENSATEMS	16941	3.6777	-245.1		PPB	28 Nov 2022 17:03:06	U	SMPL
CCV94 - 1	CCV94	23230	5.0482	-245.1		PPB	28 Nov 2022 17:05:21	U	SMPL
CCB94 - 1	CCB94	43	-0.0046	-245.1		PPB	28 Nov 2022 17:07:37	U	SMPL
N5655-01MSD - 1	SEMI-ANNUAL-CONDENSATEMSD	16692	3.6234	-245.1		PPB	28 Nov 2022 17:09:53	U	SMPL
N5655-01LX5 - 1		557	0.1074	-245.1		PPB	28 Nov 2022 17:12:08	U	SMPL
N5655-01A - 1		20262	4.4014	-245.1		PPB	28 Nov 2022 17:14:23	U	SMPL
CCV95 - 1	CCV95	21908	4.7603	-245.1		PPB	28 Nov 2022 17:16:39	U	SMPL
CCB95 - 1	CCB95	52	-0.0027	-245.1		PPB	28 Nov 2022 17:18:55	U	SMPL