



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB123047

Slope : 98.9

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	11/28/2022	16:30
2	CAL2	1	Water	NA	NA	20.8	7.00	11/28/2022	16:31
3	CAL3	1	Water	NA	NA	20.1	10.03	11/28/2022	16:32
4	ICV	1	Water	NA	NA	20.08	7.00	11/28/2022	16:33
5	CCV1	1	Water	NA	NA	20.7	2.01	11/28/2022	16:35
6	N5780-02	1	Solid	20.02	20	20.6	9.01	11/28/2022	16:38
7	N5795-02	1	Solid	20.03	20	21.4	7.89	11/28/2022	16:40
8	N5795-05	1	Solid	20.02	20	21.3	9.21	11/28/2022	16:43
9	N5796-02	1	Solid	20.03	20	21.9	7.83	11/28/2022	16:45
10	N5796-05	1	Solid	20.02	20	21.6	7.66	11/28/2022	16:46
11	N5796-05DUP	1	Solid	20.03	20	21.7	7.67	11/28/2022	16:47
12	CCV2	1	Water	NA	NA	20.9	12.03	11/28/2022	16:49

UB 123047

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity n5780 WorkList ID : 165345 Department : Wet-Chemistry Date : 11-28-2022 12:42:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/23/2022	Solid	N5780-02	Corrosivity	Cool 4 deg C	PSEG03	D11	TP-M	11/23/2022	9045D
11/28/2022	Solid	N5791-02	Corrosivity	Cool 4 deg C	NEWP02	M11	WSH-STOCK-1128	11/28/2022	9045D
11/28/2022	Solid	N5795-02	Corrosivity	Cool 4 deg C	PSEG03	M11	TP-10	11/28/2022	9045D
11/28/2022	Solid	N5795-05	Corrosivity	Cool 4 deg C	PSEG03	M11	TP-12	11/28/2022	9045D
11/28/2022	Solid	N5796-02	Corrosivity	Cool 4 deg C	PSEG03	M11	MH-15	11/28/2022	9045D
11/28/2022	Solid	N5796-05	Corrosivity	Cool 4 deg C	PSEG03	M11	TP-N	11/28/2022	9045D

Date/Time 11.28.22 16:20
 Raw Sample Received by: 201 WCC
 Raw Sample Relinquished by: CD SM

Date/Time 11.28.22 18:00
 Raw Sample Received by: CD SM
 Raw Sample Relinquished by: 201 WCC