



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB123052

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	11/29/2022	12:25
2	CAL2	1	Water	NA	NA	20.7	7.01	11/29/2022	12:26
3	CAL3	1	Water	NA	NA	20.7	10.03	11/29/2022	12:27
4	ICV	1	Water	NA	NA	20.8	6.98	11/29/2022	12:30
5	CCV1	1	Water	NA	NA	20.7	2.01	11/29/2022	12:33
6	N5800-01	1	Water	NA	NA	20.1	5.27	11/29/2022	12:36
7	N5800-01DUP	1	Water	NA	NA	20.1	5.28	11/29/2022	12:37
8	N5800-02	1	Water	NA	NA	20.2	5.42	11/29/2022	12:40
9	CCV2	1	Water	NA	NA	20.8	12.03	11/29/2022	12:42

WORKLIST(Hardcopy Internal Chain)

WB 23052

WorkList Name : PH N5800

WorkList ID : 165370

Department : Wet-Chemistry

Date : 11-29-2022 12:16:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/28/2022	Water	N5800-01	pH	Cool 4 deg C	TETR06	M11	RW9-PT-IN01-11282022	11/28/2022	9040C
11/28/2022	Water	N5800-02	pH	Cool 4 deg C	TETR06	M11	RW9-PT-OUT01-11282022	11/28/2022	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

11.29.22 12:20

201000

201000

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

11.29.22

201000

201000