



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB123053

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	11/29/2022	12:25
2	CAL2	1	Water	NA	NA	20.7	7.01	11/29/2022	12:26
3	CAL3	1	Water	NA	NA	20.7	10.03	11/29/2022	12:27
4	ICV	1	Water	NA	NA	20.8	6.98	11/29/2022	12:30
5	CCV1	1	Water	NA	NA	20.6	2.01	11/29/2022	12:45
6	N5791-01	1	Solid	20.02	20	21.7	7.36	11/29/2022	12:50
7	N5791-01DUP	1	Solid	20.03	20	21.8	7.37	11/29/2022	12:51
8	CCV2	1	Water	NA	NA	20.9	12.03	11/29/2022	13:00

113 123053

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph s n5791 WorkList ID : 165371 Department : Wet-Chemistry Date : 11-29-2022 12:18:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/28/2022	Solid	N5791-01	pH	Cool 4 deg C	NEWPO2	M11	WSH-STOCK-1128	11/28/2022	9045D

Date/Time 11-29-22 12:20
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 11-29-22 13:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]