

LB123060


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

43192



A= 0.0000e+000

B= 2.3213e-004

C= -2.6160e-002

Rho= 0.9998979

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Yol
0.0	0.000	-0.013	-0.013	56	0.000	56					-6
0.2	0.200	0.188	-0.012	924	0.0 %	924					-1
2.5	2.500	2.487	-0.013	10825	0.0 %	10825					2
5.0	5.000	5.107	0.107	22113	0.0 %	22113					-1
7.5	7.500	7.431	-0.069	32125	0.0 %	32125					0
10.0	10.000	10.000	0.000	43192	0.0 %	43192					

LB123060 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Nov 2022 08:18:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	56	-	0.0000	7470A	PPB	30 Nov 2022 08:24:34	S	Std
0.2 - 1	SD-2	924	-	0.2000	7470A	PPB	30 Nov 2022 08:26:51	S	Std
2.5 - 1	SD-5	10825	-	2.5000	7470A	PPB	30 Nov 2022 08:29:09	S	Std
5.0 - 1	SS	22113	-	5.0000	7470A	PPB	30 Nov 2022 08:31:25	S	Std
7.5 - 1	SD-5	32125	-	7.5000	7470A	PPB	30 Nov 2022 08:33:42	S	Std
10.0 - 1	SD	43192	-	10.0000	7470A	PPB	30 Nov 2022 08:35:58	S	Std
ICV39 - 1	ICV39	17668	4.0752	-	7470A	PPB	30 Nov 2022 08:38:46	U	SMPL
ICB39 - 1	ICB39	118	0.0012	-	7470A	PPB	30 Nov 2022 08:41:00	U	SMPL
CCV01 - 1	CCV01	22228	5.1337	-	7470A	PPB	30 Nov 2022 08:43:16	U	SMPL
CCB01 - 1	CCB01	-28	-0.0327	-	7470A	PPB	30 Nov 2022 08:45:31	U	SMPL
CRA - 1	CRA	907	0.1844	-	7470A	PPB	30 Nov 2022 08:47:46	U	SMPL
HighStd - 1	HighStd	43189	9.9994	-	7470A	PPB	30 Nov 2022 08:50:01	U	SMPL
ChkStd - 1	ChkStd	31789	7.3531	-	7470A	PPB	30 Nov 2022 08:52:16	U	SMPL
PB149336BL - 1	PBW	-9	-0.0282	-	7470A	PPB	30 Nov 2022 08:54:32	U	SMPL
PB149336BS - 1	LCSW	17522	4.0413	-	7470A	PPB	30 Nov 2022 08:56:49	U	SMPL
N5780-02 - 1	TP-M	12	-0.0234	-	7470A	PPB	30 Nov 2022 08:59:05	U	SMPL
N5782-02 - 1	CONCRETE-PILE	23	-0.0208	-	7470A	PPB	30 Nov 2022 09:01:22	U	SMPL
N5789-03 - 1	TFI-359	-3	-0.0269	-	7470A	PPB	30 Nov 2022 09:03:39	U	SMPL
N5789-06 - 1	SP-1	31	-0.0190	-	7470A	PPB	30 Nov 2022 09:05:56	U	SMPL
N5793-02 - 1	B-40-SB01	-17	-0.0301	-	7470A	PPB	30 Nov 2022 09:08:12	U	SMPL
CCV02 - 1	CCV02	22045	5.0912	-	7470A	PPB	30 Nov 2022 09:10:26	U	SMPL
CCB02 - 1	CCB02	-22	-0.0313	-	7470A	PPB	30 Nov 2022 09:12:42	U	SMPL
N5795-02 - 1	TP-10	70	-0.0099	-	7470A	PPB	30 Nov 2022 09:14:57	U	SMPL
N5795-05 - 1	TP-12	74	-0.0090	-	7470A	PPB	30 Nov 2022 09:17:13	U	SMPL
N5796-02 - 1	MH-15	56	-0.0132	-	7470A	PPB	30 Nov 2022 09:19:29	U	SMPL
N5796-05 - 1	TP-N	112	-0.0002	-	7470A	PPB	30 Nov 2022 09:21:45	U	SMPL
N5796-05DUP - 1	TP-NDUP	61	-0.0120	-	7470A	PPB	30 Nov 2022 09:24:01	U	SMPL
N5796-05MS - 1	TP-NMS	16738	3.8593	-	7470A	PPB	30 Nov 2022 09:26:17	U	SMPL
N5796-05MSD - 1	TP-NMSD	17613	4.0624	-	7470A	PPB	30 Nov 2022 09:28:33	U	SMPL
N5796-05LX5 - 1		-22	-0.0313	-	7470A	PPB	30 Nov 2022 09:30:50	U	SMPL
N5796-05A - 1		17563	4.0508	-	7470A	PPB	30 Nov 2022 09:33:07	U	SMPL
PB149313TB - 1	PB149313TB	20	-0.0215	-	7470A	PPB	30 Nov 2022 09:35:23	U	SMPL
CCV03 - 1	CCV03	22381	5.1692	-	7470A	PPB	30 Nov 2022 09:37:38	U	SMPL
CCB03 - 1	CCB03	-26	-0.0322	-	7470A	PPB	30 Nov 2022 09:39:54	U	SMPL