



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123061

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	12/01/2022	09:00
2	N5805-01		1	13	100.00	>212.0	>212.0	12/01/2022	09:30
3	N5805-01DUP		1	14	100.00	>212.0	>212.0	12/01/2022	10:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

26123061

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp11-22

WorkList ID : 165252

Department : Wet-Chemistry

Date : 11-22-2022 08:37:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/06/2022	Water	N5805-01	Flash Point	Cool 4 deg C	PSEG03	M11	NWB-1861	11/29/2022	1010B

Date/Time 12/11/2022 08:45
 Raw Sample Received by: RIM CMC2
 Raw Sample Relinquished by: CP(SM)

Date/Time 12/11/2022 10:45
 Raw Sample Received by: CP(SM)
 Raw Sample Relinquished by: RIM CMC2