

LB123062


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

52332



A= 0.0000e+000

B= 1.9003e-004

C= 1.2390e-003

Rho= 0.9999021

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/20
0.0	0.000	0.006	0.006	23	0.000	23					
0.2	0.200	0.179	-0.021	934	0.0 %	934					-1)
2.5	2.500	2.524	0.024	13277	0.0 %	13277					-1)
5.0	5.000	4.949	-0.051	26037	0.0 %	26037					-1)
7.5	7.500	7.597	0.097	39971	0.0 %	39971					-1)
10.0	10.000	9.946	-0.054	52332	0.0 %	52332					-1)

LB123062 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 30 Nov 2022 10:23:30

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	23	-	0.0000	7471B	PPB	30 Nov 2022 10:29:53	S	Std
0.2 - 1	50.2	934	-	0.2000	7471B	PPB	30 Nov 2022 10:32:12	S	Std
2.5 - 1	52.5	13277	-	2.5000	7471B	PPB	30 Nov 2022 10:34:32	S	Std
5.0 - 1	55	26037	-	5.0000	7471B	PPB	30 Nov 2022 10:36:49	S	Std
7.5 - 1	57.5	39971	-	7.5000	7471B	PPB	30 Nov 2022 10:39:05	S	Std
10.0 - 1	510	52332	-	10.0000	7471B	PPB	30 Nov 2022 10:41:21	S	Std
ICV40 - 1	ICV40	20866	3.9663	-	7471B	PPB	30 Nov 2022 10:44:02	U	SMPL
ICB40 - 1	ICB40	165	0.0326	-	7471B	PPB	30 Nov 2022 10:46:17	U	SMPL
CCV04 - 1	CCV04	26248	4.9891	-	7471B	PPB	30 Nov 2022 10:48:32	U	SMPL
CCB04 - 1	CCB04	43	0.0094	-	7471B	PPB	30 Nov 2022 10:50:47	U	SMPL
CRA - 1	CRA	982	0.1878	-	7471B	PPB	30 Nov 2022 10:53:02	U	SMPL
HighStd - 1	HighStd	51978	9.8784	-	7471B	PPB	30 Nov 2022 10:55:17	U	SMPL
ChkStd - 1	ChkStd	37725	7.1700	-	7471B	PPB	30 Nov 2022 10:57:32	U	SMPL
PB149342BL - 1	PBS	0	0.0012	-	7471B	PPB	30 Nov 2022 10:59:48	U	SMPL
PB149342BS - 1	LCSS	18820	3.5775	-	7471B	PPB	30 Nov 2022 11:02:04	U	SMPL
N5782-01 - 1	CONCRETE-PILE	583	0.1120	-	7471B	PPB	30 Nov 2022 11:04:21	U	SMPL
N5787-01 - 1	SB-24-STOCK	2839	0.5407	-	7471B	PPB	30 Nov 2022 11:06:37	U	SMPL
N5789-01 - 1	TFI-359	5648	1.0745	-	7471B	PPB	30 Nov 2022 11:08:53	U	SMPL
N5789-04 - 1	SP-1	8136	1.5473	-	7471B	PPB	30 Nov 2022 11:11:11	U	SMPL
N5791-01 - 1	WSH-STOCK-1128	4883	0.9291	-	7471B	PPB	30 Nov 2022 11:13:28	U	SMPL
CCV05 - 1	CCV05	25534	4.8534	-	7471B	PPB	30 Nov 2022 11:15:44	U	SMPL
CCB05 - 1	CCB05	-87	-0.0153	-	7471B	PPB	30 Nov 2022 11:17:59	U	SMPL
N5792-01 - 1	OR-02-112822	1701	0.3245	-	7471B	PPB	30 Nov 2022 11:20:14	U	SMPL
N5793-01 - 1	B-40-SB01	25122	4.7751	-	7471B	PPB	30 Nov 2022 11:22:29	U	SMPL
N5793-03 - 1	B-40-SB02	230	0.0449	-	7471B	PPB	30 Nov 2022 11:24:45	U	SMPL
N5795-01 - 1	TP-10	8507	1.6178	-	7471B	PPB	30 Nov 2022 11:27:01	U	SMPL
N5795-04 - 1	TP-12	5938	1.1296	-	7471B	PPB	30 Nov 2022 11:29:17	U	SMPL
N5796-01 - 1	MH-15	1896	0.3615	-	7471B	PPB	30 Nov 2022 11:31:33	U	SMPL
N5796-01DUP - 1	MH-15DUP	2041	0.3891	-	7471B	PPB	30 Nov 2022 11:33:53	U	SMPL
N5796-01MS - 1	MH-15MS	20735	3.9414	-	7471B	PPB	30 Nov 2022 11:36:10	U	SMPL
N5796-01MSD - 1	MH-15MSD	20797	3.9532	-	7471B	PPB	30 Nov 2022 11:38:27	U	SMPL
N5796-04 - 1	TP-N	266	0.0518	-	7471B	PPB	30 Nov 2022 11:40:43	U	SMPL
CCV06 - 1	CCV06	25918	4.9263	-	7471B	PPB	30 Nov 2022 11:42:58	U	SMPL
CCB06 - 1	CCB06	-22	-0.0029	-	7471B	PPB	30 Nov 2022 11:45:14	U	SMPL
N5801-01 - 1	TP-13	5233	0.9956	-	7471B	PPB	30 Nov 2022 11:47:29	U	SMPL
N5812-01 - 1	MH-16	3237	0.6164	-	7471B	PPB	30 Nov 2022 11:49:45	U	SMPL
N5796-01LX5 - 1		276	0.0537	-	7471B	PPB	30 Nov 2022 11:52:01	U	SMPL
N5796-01A - 1		21862	4.1556	-	7471B	PPB	30 Nov 2022 11:54:16	U	SMPL
CCV07 - 1	CCV07	25841	4.9117	-	7471B	PPB	30 Nov 2022 12:08:02	U	SMPL
CCB07 - 1	CCB07	-159	-0.0290	-	7471B	PPB	30 Nov 2022 12:10:18	U	SMPL