

LB123073


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

55853



A= 0.0000e+000

B= 1.7822e-004

C= 1.2684e-003

Rho= 0.9999337

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.001	-0.001	-10	0.000	-10					
0.2	0.200	0.178	-0.022	990	0.0 %	990					-11
2.5	2.500	2.470	-0.030	13854	0.0 %	13854					-1
5.0	5.000	5.086	0.086	28529	0.0 %	28529					2
7.5	7.500	7.511	0.011	42139	0.0 %	42139					0
10.0	10.000	9.955	-0.045	55853	0.0 %	55853					0

LB123073 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 01 Dec 2022 12:26:32

Sample ID	Extended ID	μ Abs.	Conc.	Strd Conc	Method	Units	Date	Type	Type
0.0 - 1	SV	-10	-	0.0000	7471B	PPB	01 Dec 2022 12:32:53	S	Std
0.2 - 1	SV	990	-	0.2000	7471B	PPB	01 Dec 2022 12:35:11	S	Std
2.5 - 1	SV	13854	-	2.5000	7471B	PPB	01 Dec 2022 12:37:29	S	Std
5.0 - 1	SV	28529	-	5.0000	7471B	PPB	01 Dec 2022 12:39:45	S	Std
7.5 - 1	SV	42139	-	7.5000	7471B	PPB	01 Dec 2022 12:42:02	S	Std
10.0 - 1	SV	55853	-	10.0000	7471B	PPB	01 Dec 2022 12:44:18	S	Std
ICV41 - 1	ICV41	21070	3.7564	-7471B	PPB	PPB	01 Dec 2022 12:47:00	U	SMPL
ICB41 - 1	ICB41	51	0.0104	-7471B	PPB	PPB	01 Dec 2022 12:49:15	U	SMPL
CCV08 - 1	CCV08	28750	5.1251	-7471B	PPB	PPB	01 Dec 2022 12:51:30	U	SMPL
CCB08 - 1	CCB08	-62	-0.0098	-7471B	PPB	PPB	01 Dec 2022 12:53:45	U	SMPL
CRA - 1	CRA	959	0.1722	-7471B	PPB	PPB	01 Dec 2022 12:56:00	U	SMPL
HighStd - 1	HighStd	55316	9.8597	-7471B	PPB	PPB	01 Dec 2022 12:58:15	U	SMPL
ChkStd - 1	ChkStd	39669	7.0711	-7471B	PPB	PPB	01 Dec 2022 13:00:31	U	SMPL
PB149351BL - 1	PBS	-104	-0.0173	-7471B	PPB	PPB	01 Dec 2022 13:02:47	U	SMPL
PB149351BS - 1	LCSS	21834	3.8925	-7471B	PPB	PPB	01 Dec 2022 13:05:03	U	SMPL
N5799-01 - 1	220-1-STONES	1039	0.1864	-7471B	PPB	PPB	01 Dec 2022 13:07:19	U	SMPL
N5803-01 - 1	PEQ-STOCK	3086	0.5513	-7471B	PPB	PPB	01 Dec 2022 13:09:35	U	SMPL
N5804-01 - 1	SB-1-STOCK	18215	3.2476	-7471B	PPB	PPB	01 Dec 2022 13:11:52	U	SMPL
N5806-01 - 1	COMP-1	3930	0.7017	-7471B	PPB	PPB	01 Dec 2022 13:14:09	U	SMPL
N5806-08 - 1	COMP-2	4381	0.7821	-7471B	PPB	PPB	01 Dec 2022 13:16:25	U	SMPL
CCV09 - 1	CCV09	28570	5.0930	-7471B	PPB	PPB	01 Dec 2022 13:18:41	U	SMPL
CCB09 - 1	CCB09	-14	-0.0012	-7471B	PPB	PPB	01 Dec 2022 13:20:56	U	SMPL
N5807-04 - 1	COMP-3	2886	0.5156	-7471B	PPB	PPB	01 Dec 2022 13:23:11	U	SMPL
N5809-01 - 1	HOWE-DRUM-A-1129	6426	1.1465	-7471B	PPB	PPB	01 Dec 2022 13:25:26	U	SMPL
N5809-03 - 1	HOWE-DRUM-B-1129	1073	0.1925	-7471B	PPB	PPB	01 Dec 2022 13:27:41	U	SMPL
N5809-05 - 1	HOWE-DRUM-C-1129	1020	0.1831	-7471B	PPB	PPB	01 Dec 2022 13:29:57	U	SMPL
N5809-05DUP - 1	HOWE-DRUM-C-1129DUP	1089	0.1954	-7471B	PPB	PPB	01 Dec 2022 13:32:13	U	SMPL
N5809-05MS - 1	HOWE-DRUM-C-1129MS	22489	4.0093	-7471B	PPB	PPB	01 Dec 2022 13:34:30	U	SMPL
N5809-05MSD - 1	HOWE-DRUM-C-1129MSD	22031	3.9276	-7471B	PPB	PPB	01 Dec 2022 13:36:47	U	SMPL
N5809-05LX5 - 1		0	0.0013	-7471B	PPB	PPB	01 Dec 2022 13:39:04	U	SMPL
N5809-05LX5 - 1		301	0.0549	-7471B	PPB	PPB	01 Dec 2022 13:41:25	U	SMPL
N5809-05A - 1		22302	3.9759	-7471B	PPB	PPB	01 Dec 2022 13:43:42	U	SMPL
CCV10 - 1	CCV10	28154	5.0189	-7471B	PPB	PPB	01 Dec 2022 13:45:57	U	SMPL
CCB10 - 1	CCB10	-129	-0.0217	-7471B	PPB	PPB	01 Dec 2022 13:48:13	U	SMPL