

LB123075


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

48069



A= 0.0000e+000

B= 2.0789e-004

C= 4.6750e-002

Rho= 0.9995500

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.039	0.039	-37	0.000	-37					
0.2	0.200	0.204	0.004	755	0.0 %	755					
2.5	2.500	2.378	-0.122	11214	0.0 %	11214					
5.0	5.000	5.185	0.185	24717	0.0 %	24717					
7.5	7.500	7.354	-0.146	35151	0.0 %	35151					
10.0	10.000	10.040	0.040	48069	0.0 %	48069					

1
2
-5
4
-2
0

LB123075 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 Dec 2022 14:17:25

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-37	-	0.0000	7470A	PPB	01 Dec 2022 14:17:43	S	Std
0.2 - 1	50.2	755	-	0.2000	7470A	PPB	01 Dec 2022 14:20:00	S	Std
2.5 - 1	52.5	11214	-	2.5000	7470A	PPB	01 Dec 2022 14:22:18	S	Std
5.0 - 1	55	24717	-	5.0000	7470A	PPB	01 Dec 2022 14:24:35	S	Std
7.5 - 1	57.5	35151	-	7.5000	7470A	PPB	01 Dec 2022 14:26:52	S	Std
10.0 - 1	510	48069	-	10.0000	7470A	PPB	01 Dec 2022 14:29:08	S	Std
ICV42 - 1	ICV42	17518	3.6886	-	7470A	PPB	01 Dec 2022 14:31:53	U	SMPL
ICB42 - 1	ICB42	109	0.0694	-	7470A	PPB	01 Dec 2022 14:34:08	U	SMPL
CCV11 - 1	CCV11	23910	5.0174	-	7470A	PPB	01 Dec 2022 14:41:33	U	SMPL
CCB11 - 1	CCB11	-52	0.0359	-	7470A	PPB	01 Dec 2022 14:43:48	U	SMPL
CRA - 1	CRA	836	0.2205	-	7470A	PPB	01 Dec 2022 14:46:03	U	SMPL
HighStd - 1	HighStd	46088	9.6280	-	7470A	PPB	01 Dec 2022 14:48:18	U	SMPL
ChkStd - 1	ChkStd	33323	6.9742	-	7470A	PPB	01 Dec 2022 14:50:33	U	SMPL
PB149352BL - 1	PBW	-2	0.0463	-	7470A	PPB	01 Dec 2022 14:52:49	U	SMPL
PB149352BS - 1	LCSW	17562	3.6977	-	7470A	PPB	01 Dec 2022 14:55:04	U	SMPL
N5799-03 - 1	220-1-STONES	-25	0.0416	-	7470A	PPB	01 Dec 2022 14:57:20	U	SMPL
N5801-02 - 1	TP-13	-11	0.0445	-	7470A	PPB	01 Dec 2022 14:59:36	U	SMPL
N5806-02 - 1	COMP-1	-51	0.0361	-	7470A	PPB	01 Dec 2022 15:01:56	U	SMPL
N5806-09 - 1	COMP-2	-54	0.0355	-	7470A	PPB	01 Dec 2022 15:04:13	U	SMPL
N5807-05 - 1	COMP-3	-54	0.0355	-	7470A	PPB	01 Dec 2022 15:06:29	U	SMPL
CCV12 - 1	CCV12	23424	4.9164	-	7470A	PPB	01 Dec 2022 15:08:44	U	SMPL
CCB12 - 1	CCB12	-24	0.0418	-	7470A	PPB	01 Dec 2022 15:10:59	U	SMPL
N5809-02 - 1	HOWE-DRUM-A-1129	85	0.0644	-	7470A	PPB	01 Dec 2022 15:13:14	U	SMPL
N5809-04 - 1	HOWE-DRUM-B-1129	92	0.0659	-	7470A	PPB	01 Dec 2022 15:15:30	U	SMPL
N5809-06 - 1	HOWE-DRUM-C-1129	53	0.0578	-	7470A	PPB	01 Dec 2022 15:17:46	U	SMPL
N5812-02 - 1	MH-16	47	0.0565	-	7470A	PPB	01 Dec 2022 15:20:01	U	SMPL
N5812-02DUP - 1	MH-16DUP	21	0.0511	-	7470A	PPB	01 Dec 2022 15:22:17	U	SMPL
N5812-02MS - 1	MH-16MS	16218	3.4183	-	7470A	PPB	01 Dec 2022 15:24:33	U	SMPL
N5812-02MSD - 1	MH-16MSD	16771	3.5333	-	7470A	PPB	01 Dec 2022 15:26:50	U	SMPL
N5812-02LX5 - 1		-32	0.0401	-	7470A	PPB	01 Dec 2022 15:29:07	U	SMPL
N5812-02A - 1		17839	3.7553	-	7470A	PPB	01 Dec 2022 15:31:24	U	SMPL
PB149340TB - 1	PB149340TB	-89	0.0282	-	7470A	PPB	01 Dec 2022 15:33:40	U	SMPL
CCV13 - 1	CCV13	24120	5.0610	-	7470A	PPB	01 Dec 2022 15:36:03	U	SMPL
CCB13 - 1	CCB13	-34	0.0397	-	7470A	PPB	01 Dec 2022 15:38:19	U	SMPL