

LB123079


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

47502



A= 0.0000e+000

B= 2.1037e-004 slope

C= 2.6324e-002 y-intercept

Rho= 0.9999672

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.027	0.027	3	0.000	3					
0.05	0.050					185	0				
0.20	0.200	0.207	0.007	860	0.0 %	860					
2.50	2.500	2.491	-0.009	11714	0.0 %	11714					
5.00	5.000	4.938	-0.062	23349	0.0 %	23349					
7.50	7.500	7.517	0.017	35609	0.0 %	35609					
10.0	10.000	10.019	0.019	47502	0.0 %	47502					

0.0-10.0

LB123079

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0		3 -		0 SFAM01.1	PPB	12/2/2022 8:25	Std
	0.2 S01	860	-		0.2 SFAM01.1	PPB	12/2/2022 8:30	Std
	2.5 S02	11714	-		2.5 SFAM01.1	PPB	12/2/2022 8:32	Std
	5 S03	23349	-		5 SFAM01.1	PPB	12/2/2022 8:34	Std
	7.5 S04	35609	-		7.5 SFAM01.1	PPB	12/2/2022 8:37	Std
	10 S05	47502	-		10 SFAM01.1	PPB	12/2/2022 8:39	Std
ICV043	ICV043	18199	3.8549 -		SFAM01.1	PPB	12/2/2022 8:42	SMPL
ICB043	ICB043	111	0.0497 -		SFAM01.1	PPB	12/2/2022 8:44	SMPL
CCV014	CCV014	23196	4.9061 -		SFAM01.1	PPB	12/2/2022 8:46	SMPL
CCB014	CCB014	-85	0.0084 -		SFAM01.1	PPB	12/2/2022 8:49	SMPL
PB149357BL	PBS357	-10	0.0242 -		SFAM01.1	PPB	12/2/2022 8:51	SMPL
N5631-01	YBK66	1245381	HIGH -		SFAM01.1	PPB	12/2/2022 8:53	
N5631-02	YBK66D	1247782	HIGH -		SFAM01.1	PPB	12/2/2022 8:55	
N5631-03	YBK66S	1233969	HIGH -		SFAM01.1	PPB	12/2/2022 8:57	
N5654-01	COAA0	6748	1.4459 -		SFAM01.1	PPB	12/2/2022 9:08	SMPL
N5654-02	COAA1	7262	1.554 -		SFAM01.1	PPB	12/2/2022 9:10	SMPL
N5654-03	COAA2	1559	0.3543 -		SFAM01.1	PPB	12/2/2022 9:13	SMPL
N5654-04	COAA3	2449	0.5415 -		SFAM01.1	PPB	12/2/2022 9:15	SMPL
N5654-05	COAA4	9045	1.9291 -		SFAM01.1	PPB	12/2/2022 9:17	SMPL
N5654-06	COAA5	7917	1.6918 -		SFAM01.1	PPB	12/2/2022 9:20	SMPL
N5654-07	COAA6	2780	0.6112 -		SFAM01.1	PPB	12/2/2022 9:22	SMPL
N5654-08	COAA6D	3137	0.6863 -		SFAM01.1	PPB	12/2/2022 9:24	SMPL
N5654-09	COAA6S	13962	2.9635 -		SFAM01.1	PPB	12/2/2022 9:26	SMPL
N5654-10	COAA8	4269	0.9244 -		SFAM01.1	PPB	12/2/2022 9:29	SMPL
N5654-11	COAA9	4262	0.9229 -		SFAM01.1	PPB	12/2/2022 9:31	SMPL
PB149359BL	PBS359	-10	0.0242 -		SFAM01.1	PPB	12/2/2022 9:33	SMPL
N5657-01	COAA7	3312	0.7231 -		SFAM01.1	PPB	12/2/2022 9:35	SMPL
N5657-02	COAA7D	3386	0.7386 -		SFAM01.1	PPB	12/2/2022 9:38	SMPL
N5657-03	COAA7S	14611	3.1001 -		SFAM01.1	PPB	12/2/2022 9:40	SMPL
CCV015	CCV015	23167	4.9 -		SFAM01.1	PPB	12/2/2022 9:42	SMPL
CCB015	CCB015	-31	0.0198 -		SFAM01.1	PPB	12/2/2022 9:45	SMPL
N5657-04	COAB0	7239	1.5492 -		SFAM01.1	PPB	12/2/2022 9:47	SMPL
N5657-05	COAB1	5478	1.1787 -		SFAM01.1	PPB	12/2/2022 9:49	SMPL
N5657-06	COAB2	3244	0.7088 -		SFAM01.1	PPB	12/2/2022 9:51	SMPL
N5657-07	COAB3	3450	0.7521 -		SFAM01.1	PPB	12/2/2022 9:54	SMPL
N5657-08	COAB4	2679	0.5899 -		SFAM01.1	PPB	12/2/2022 9:56	SMPL
N5657-09	COAB5	3075	0.6732 -		SFAM01.1	PPB	12/2/2022 9:58	SMPL
N5657-10	COAB6	2565	0.5659 -		SFAM01.1	PPB	12/2/2022 10:00	SMPL
N5657-11	COAB7	2472	0.5464 -		SFAM01.1	PPB	12/2/2022 10:03	SMPL
N5657-12	COAB8	2377	0.5264 -		SFAM01.1	PPB	12/2/2022 10:05	SMPL
N5657-13	COAB9	2369	0.5247 -		SFAM01.1	PPB	12/2/2022 10:07	SMPL
N5631-01DLX10	YBK66	288158	60.6467 -		SFAM01.1	PPB	12/2/2022 10:09	SMPL
N5631-02DLX10	YBK66D	288426	60.7031 -		SFAM01.1	PPB	12/2/2022 10:12	SMPL
N5631-03DLX10	YBK66S	288079	60.6301 -		SFAM01.1	PPB	12/2/2022 10:14	SMPL
PB149358BL	PBW358	172	0.0625 -		SFAM01.1	PPB	12/2/2022 10:21	SMPL
CCV016	CCV016	22321	4.722 -		SFAM01.1	PPB	12/2/2022 10:23	SMPL

LB123079

INSTRUMENT ID : CV1

CCB016	CCB016	-557	-0.0909 -	SFAM01.1	PPB	12/2/2022 10:28 SMPL
N5631-01DL2X100	YBK66	25504	5.3917 -	SFAM01.1	PPB	12/2/2022 10:30 SMPL
N5631-02DL2X100	YBK66D	25977	5.4912 -	SFAM01.1	PPB	12/2/2022 10:42 SMPL
N5631-03DL2X100	YBK66S	28374	5.9954 -	SFAM01.1	PPB	12/2/2022 10:44 SMPL
CCV017	CCV017	23415	4.9522 -	SFAM01.1	PPB	12/2/2022 10:49 SMPL
CCB017	CCB017	-56	0.0145 -	SFAM01.1	PPB	12/2/2022 10:51 SMPL