



TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 104 °C 12/01/2022 17:10 TEMP1 OUT: 104 °C 12/02/2022 08:00
TEMP2 IN: 104 °C 12/02/2022 08:30 TEMP2 OUT: 104 °C 12/02/2022 10:00
TEMP3 IN: 550 °C 12/02/2022 10:30 TEMP3 OUT: 550 °C 12/02/2022 12:00
TEMP4 IN: 560 °C 12/02/2022 12:30 TEMP4 OUT: 560 °C 12/02/2022 14:01

Run Number: LB123081

SUPERVISOR: Iwona

ANALYST: jignesh

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB123081BL	75.9681	75.9681	100	75.9682	75.9682	75.9682	75.9682	0.0000	0
2	N5830-01	88.9401	88.9401	100	89.0105	89.0105	88.9725	88.9725	0.0380	380
3	N5830-01DUP	91.5861	91.5861	100	91.6563	91.6563	91.6194	91.6194	0.0369	369

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105 °C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105 °C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$

WORKLIST(Hardcopy Internal Chain)

VB 123081

WorkList Name : TVS N4830 W

WorkList ID : 165427

Department : Wet-Chemistry

Date : 12-01-2022 16:43:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/08/2022	Water	N5830-01	TVS	Cool 4 deg C	PSEG03	M11	COMP1-1201	12/01/2022	SM2540 G

Date/Time 12/01/22 16:50
 Raw Sample Received by: JD/CC
 Raw Sample Relinquished by: JD/CC

Date/Time 12/01/22 18:00
 Raw Sample Received by: JD/CC
 Raw Sample Relinquished by: JD/CC