



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB123082

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	12/01/2022	17:50
2	CAL2	1	Water	NA	NA	20.8	7.00	12/01/2022	17:51
3	CAL3	1	Water	NA	NA	20.8	10.03	12/01/2022	17:53
4	ICV	1	Water	NA	NA	20.7	6.98	12/01/2022	18:00
5	CCV1	1	Water	NA	NA	20.7	2.01	12/01/2022	18:02
6	N5833-02	1	Solid	20.02	20	22.6	7.32	12/01/2022	18:05
7	N5844-02	1	Solid	20.03	20	23.2	7.00	12/01/2022	18:06
8	N5844-02DUP	1	Solid	20.04	20	23.2	7.01	12/01/2022	18:07
9	CCV2	1	Water	NA	NA	20.8	12.03	12/01/2022	18:10

WORKLIST(Hardcopy Internal Chain)

12/23/22

WorkList Name : CORRSOVITY N5844

WorkList ID : 165423

Department : Wet-Chemistry

Date : 12-01-2022 16:06:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/01/2022	Solid	N5833-02	Corrosivity	Cool 4 deg C	PSEG03	M11	COMPOSITE-1	12/01/2022	9045D
12/01/2022	Solid	N5844-02	Corrosivity	Cool 4 deg C	PSEG03	M11	MH-17	12/01/2022	9045D

Date/Time 12-01-22 17:40
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 12-01-22 18:40
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]