



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123088

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.7	12/02/2022	12:10
2	N5830-01		1	13	100.00	>212.0	>212.0	12/02/2022	12:40
3	N5830-01DUP		1	14	100.00	>212.0	>212.0	12/02/2022	13:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

Lb123088

WorkList Name : fp12-2

WorkList ID : 165442

Department : Wet-Chemistry

Date : 12-02-2022 08:59:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/08/2022	Water	N5830-01	Flash Point	Cool 4 deg C	PSEG03	M11	COMP1-1201	12/01/2022	1010B
12/08/2022	Water	N5838-01	Flash Point	Cool 4 deg C	PSEG03	N11	WG-1-COMP	12/01/2022	1010B

53

12/2/22

Date/Time 12/2/2022 12:00
Raw Sample Received by: RM Cwe
Raw Sample Relinquished by: CP(SM)

Date/Time 12/2/2022 13:50
Raw Sample Received by: CP(SM)
Raw Sample Relinquished by: RM Cwe