



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123089

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.7	12/02/2022	10:20
2	N5838-03		1	12	100.00	>212.0	>212.0	12/02/2022	10:50
3	N5838-03DUP		1	14	100.00	>212.0	>212.0	12/02/2022	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66123089

WorkList Name : fp12-

WorkList ID : 165463

Department : Wet-Chemistry

Date : 12-02-2022 13:39:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/08/2022	Solid	N5838-03	Flash Point	Cool 4 deg C	PSEG03	N11	WC-1-COMP	12/01/2022	1010B

Date/Time 12/12/2022 10:15
 Raw Sample Received by: RMA Cwe2
 Raw Sample Relinquished by: CP(SM)

Date/Time 12/12/2022 11:55
 Raw Sample Received by: CP(SM)
 Raw Sample Relinquished by: RMA Cwe2