



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB123100

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	12/02/2022	16:00
2	CAL2	1	Water	NA	NA	20.8	7.00	12/02/2022	16:01
3	CAL3	1	Water	NA	NA	20.8	10.03	12/02/2022	16:03
4	ICV	1	Water	NA	NA	20.8	6.99	12/02/2022	16:05
5	CCV1	1	Water	NA	NA	20.7	2.01	12/02/2022	16:10
6	N5850-01	1	Water	NA	NA	20.2	7.24	12/02/2022	16:11
7	N5852-02	1	Water	NA	NA	20.1	4.81	12/02/2022	16:15
8	N5852-03	1	Water	NA	NA	20.7	5.25	12/02/2022	16:20
9	N5853-01	1	Water	NA	NA	20.1	4.80	12/02/2022	16:23
10	N5857-01	1	Water	NA	NA	20.5	7.67	12/02/2022	16:25
11	N5858-01	1	Water	NA	NA	20.8	7.73	12/02/2022	16:29
12	N5861-01	1	Water	NA	NA	20.3	5.25	12/02/2022	16:30
13	N5869-01	1	Water	NA	NA	20.7	5.94	12/02/2022	16:33
14	N5869-01DUP	1	Water	NA	NA	20.8	5.95	12/02/2022	16:34
15	CCV2	1	Water	NA	NA	20.8	12.02	12/02/2022	16:35

WORKLIST(Hardcopy Internal Chain)

123100

WorkList Name : ph n5850

WorkList ID : 165464

Department : Wet-Chemistry

Date : 12-02-2022 13:41:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/02/2022	Water	N5850-01	pH	Cool 4 deg C	PSEG03	M11	PURGE-WATER	12/02/2022	9040C
12/01/2022	Water	N5852-02	pH	Cool 4 deg C	TETR06	M62	RW9-PT-IN-03-12012022	12/01/2022	9040C
12/01/2022	Water	N5852-03	pH	Cool 4 deg C	TETR06	M62	RW9-PT-OUT-03-12012022	12/01/2022	9040C
11/30/2022	Water	N5853-01	pH	Cool 4 deg C	STER08	M62	POOL-WATER	11/30/2022	9040C
12/01/2022	Water	N5857-01	pH	Cool 4 deg C	ENVI70	M11	COMP-SAMPLE-TANK-10-11-12-31-33	12/01/2022	9040C
12/01/2022	Water	N5858-01	pH	Cool 4 deg C	ENVI70	N11	COMP-SAMPLE-26-27-28-30	12/01/2022	9040C
11/17/2022	Water	N5861-01	pH	Cool 4 deg C	STER06	J62	POOL-WATER	11/17/2022	9040C
12/01/2022	Water	N5869-01	pH	Cool 4 deg C	STER01	M11	POOL-WATER	12/01/2022	9040C

Date/Time 12/02/22 13:50
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 12/02/22 16:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]