



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB123102

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	12/03/2022	09:40
2	CAL2	1	Water	NA	NA	20.7	7.00	12/03/2022	09:41
3	CAL3	1	Water	NA	NA	20.7	10.03	12/03/2022	09:42
4	ICV	1	Water	NA	NA	20.7	6.98	12/03/2022	09:45
5	CCV1	1	Water	NA	NA	20.7	2.01	12/03/2022	09:46
6	N5859-03	1	Water	NA	NA	22.7	7.08	12/03/2022	09:49
7	N5859-03DUP	1	Water	NA	NA	22.8	7.09	12/03/2022	09:50
8	N5859-04	1	Water	NA	NA	22.6	5.84	12/03/2022	09:55
9	N5859-05	1	Water	NA	NA	22.7	7.04	12/03/2022	10:00
10	N5859-06	1	Water	NA	NA	22.7	7.72	12/03/2022	10:02
11	N5859-07	1	Water	NA	NA	22.6	5.62	12/03/2022	10:05
12	N5859-08	1	Water	NA	NA	22.7	5.05	12/03/2022	10:10
13	N5859-09	1	Water	NA	NA	22.9	7.19	12/03/2022	10:14
14	CCV2	1	Water	NA	NA	20.8	12.03	12/03/2022	10:15

WORKLIST(Hardcopy Internal Chain)

12/03/22

WorkList Name : ph n5859

WorkList ID : 165470

Department : Wet-Chemistry

Date : 12-03-2022 09:28:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/02/2022	Water	N5859-03	pH	Cool 4 deg C	PSEG03	M11	ORA-1518	12/02/2022	9040C
12/02/2022	Water	N5859-04	pH	Cool 4 deg C	PSEG03	M11	ORA-0583	12/02/2022	9040C
12/02/2022	Water	N5859-05	pH	Cool 4 deg C	PSEG03	M11	ORA-0584	12/02/2022	9040C
12/02/2022	Water	N5859-06	pH	Cool 4 deg C	PSEG03	M11	ORA-0585	12/02/2022	9040C
12/02/2022	Water	N5859-07	pH	Cool 4 deg C	PSEG03	M11	ORA-0586	12/02/2022	9040C
12/02/2022	Water	N5859-08	pH	Cool 4 deg C	PSEG03	M11	ORA-0587	12/02/2022	9040C
12/02/2022	Water	N5859-09	pH	Cool 4 deg C	PSEG03	M11	31910	12/02/2022	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

12/03/22 09:35

12/03/22

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

12:50

12/03/22