

LB123106


7471B

INSTRUMENT ID: CVL

Linear

 μ Abs.:

47643



A= 0.0000e+000

B= 2.0915e-004

C= -3.9238e-002

Rho= 0.9999052

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%v
0.0	0.000	-0.007	-0.007	156	0.000	156					5
0.2	0.200	0.178	-0.022	1041	0.0 %	1041					-11
2.5	2.500	2.508	0.008	12179	0.0 %	12179					0
5.0	5.000	4.998	-0.002	24085	0.0 %	24085					0
7.5	7.500	7.597	0.097	36509	0.0 %	36509					1
10.0	10.000	9.925	-0.075	47643	0.0 %	47643					-1

LB123106 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Dec 2022 11:11:18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	156	-	0.0000	7471B	PPB	05 Dec 2022 11:25:40	S	Std
0.2 - 1	50.2	1041	-	0.2000	7471B	PPB	05 Dec 2022 11:27:57	S	Std
2.5 - 1	52.5	12179	-	2.5000	7471B	PPB	05 Dec 2022 11:30:14	S	Std
5.0 - 1	55	24085	-	5.0000	7471B	PPB	05 Dec 2022 11:32:32	S	Std
7.5 - 1	57.5	36509	-	7.5000	7471B	PPB	05 Dec 2022 11:34:48	S	Std
10.0 - 1	510	47643	-	10.0000	7471B	PPB	05 Dec 2022 11:37:04	S	Std
ICV44 - 1	ICV44	19908	4.1245	-	7471B	PPB	05 Dec 2022 11:42:24	U	SMPL
ICB44 - 1	ICB44	90	-0.0204	-	7471B	PPB	05 Dec 2022 11:44:39	U	SMPL
CCV18 - 1	CCV18	23731	4.9241	-	7471B	PPB	05 Dec 2022 11:46:54	U	SMPL
CCB18 - 1	CCB18	-2	-0.0397	-	7471B	PPB	05 Dec 2022 11:49:09	U	SMPL
CRA - 1	CRA	1105	0.1919	-	7471B	PPB	05 Dec 2022 11:53:55	U	SMPL
HighStd - 1	HighStd	47161	9.8245	-	7471B	PPB	05 Dec 2022 11:56:10	U	SMPL
ChkStd - 1	ChkStd	33801	7.0303	-	7471B	PPB	05 Dec 2022 11:58:25	U	SMPL
PB149402BL - 1	PBS	46	-0.0296	-	7471B	PPB	05 Dec 2022 12:00:42	U	SMPL
PB149402BS - 1	LCSS	17244	3.5674	-	7471B	PPB	05 Dec 2022 12:02:58	U	SMPL
N5814-01 - 1	NB-08-113022	2048	0.3891	-	7471B	PPB	05 Dec 2022 12:11:30	U	SMPL
N5814-01DUP - 1	NB-08-113022DUP	1996	0.3782	-	7471B	PPB	05 Dec 2022 12:13:46	U	SMPL
N5814-01MS - 1	NB-08-113022MS	19068	3.9488	-	7471B	PPB	05 Dec 2022 12:16:03	U	SMPL
N5814-01MSD - 1	NB-08-113022MSD	19775	4.0967	-	7471B	PPB	05 Dec 2022 12:18:19	U	SMPL
N5817-01 - 1	OK-3-113022	1672	0.3105	-	7471B	PPB	05 Dec 2022 12:20:35	U	SMPL
CCV19 - 1	CCV19	23317	4.8375	-	7471B	PPB	05 Dec 2022 12:22:50	U	SMPL
CCB19 - 1	CCB19	43	-0.0302	-	7471B	PPB	05 Dec 2022 12:25:05	U	SMPL
N5818-04 - 1	TR-04-113022	1963	0.3713	-	7471B	PPB	05 Dec 2022 12:27:20	U	SMPL
N5823-01 - 1	OE-2-113022	1508	0.2762	-	7471B	PPB	05 Dec 2022 12:29:35	U	SMPL
N5833-01 - 1	COMPOSITE-1	1651	0.3061	-	7471B	PPB	05 Dec 2022 12:31:51	U	SMPL
N5838-03 - 1	WC-1-COMP	73	-0.0240	-	7471B	PPB	05 Dec 2022 12:34:06	U	SMPL
N5844-01 - 1	MH-17	628	0.0921	-	7471B	PPB	05 Dec 2022 12:36:23	U	SMPL
N5845-01 - 1	OR-03-120122	1506	0.2757	-	7471B	PPB	05 Dec 2022 12:38:40	U	SMPL
N5814-01LX5 - 1		307	0.0250	-	7471B	PPB	05 Dec 2022 12:40:56	U	SMPL
N5814-01A - 1		20184	4.1823	-	7471B	PPB	05 Dec 2022 12:43:14	U	SMPL
CCV20 - 1	CCV20	23028	4.7771	-	7471B	PPB	05 Dec 2022 12:45:30	U	SMPL
CCB20 - 1	CCB20	52	-0.0284	-	7471B	PPB	05 Dec 2022 12:47:47	U	SMPL