

LB123110


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

44419



A= 0.0000e+000

B= 2.3000e-004

C= 6.0894e-002

Rho= 0.9982390

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.063	0.063	9	0.000	9					17
0.2	0.200	0.233	0.033	747	0.0 %	747					8
2.5	2.500	2.701	0.201	11480	0.0 %	11480					-6
5.0	5.000	4.705	-0.295	20193	0.0 %	20193					-4
7.5	7.500	7.220	-0.280	31128	0.0 %	31128					3
10.0	10.000	10.277	0.277	44419	0.0 %	44419					

LB123110 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Dec 2022 14:20:02

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	50	9	-	0.0000	7470A	PPB	05 Dec 2022 14:23:21	S	Std
0.2 - 1	50.2	747	-	0.2000	7470A	PPB	05 Dec 2022 14:25:38	S	Std
2.5 - 1	52.5	11480	-	2.5000	7470A	PPB	05 Dec 2022 14:27:55	S	Std
5.0 - 1	55	20193	-	5.0000	7470A	PPB	05 Dec 2022 14:30:12	S	Std
7.5 - 1	57.5	31128	-	7.5000	7470A	PPB	05 Dec 2022 14:32:29	S	Std
10.0 - 1	510	44419	-	10.0000	7470A	PPB	05 Dec 2022 14:37:29	S	Std
ICV45 - 1	ICV45	17597	4.1082	-	7470A	PPB	05 Dec 2022 14:40:24	U	SMPL
ICB45 - 1	ICB45	43	0.0708	-	7470A	PPB	05 Dec 2022 14:42:38	U	SMPL
CCV21 - 1	CCV21	21021	4.8958	-	7470A	PPB	05 Dec 2022 14:44:54	U	SMPL
CCB21 - 1	CCB21	-188	0.0177	-	7470A	PPB	05 Dec 2022 14:47:09	U	SMPL
CRA - 1	CRA	691	0.2198	-	7470A	PPB	05 Dec 2022 14:49:24	U	SMPL
HighStd - 1	HighStd	43576	10.0834	-	7470A	PPB	05 Dec 2022 14:51:39	U	SMPL
ChkStd - 1	ChkStd	27611	6.4115	-	7470A	PPB	05 Dec 2022 14:53:55	U	SMPL
PB149400BL - 1	PBW	-10	0.0586	-	7470A	PPB	05 Dec 2022 14:56:11	U	SMPL
PB149400BS - 1	LCSW	17053	3.9831	-	7470A	PPB	05 Dec 2022 14:58:26	U	SMPL
N5833-02 - 1	COMPOSITE-1	-18	0.0568	-	7470A	PPB	05 Dec 2022 15:00:43	U	SMPL
N5844-02 - 1	MH-17	-15	0.0574	-	7470A	PPB	05 Dec 2022 15:02:59	U	SMPL
N5844-02DUP - 1	MH-17DUP	-33	0.0533	-	7470A	PPB	05 Dec 2022 15:05:16	U	SMPL
N5844-02MS - 1	MH-17MS	16056	3.7538	-	7470A	PPB	05 Dec 2022 15:07:33	U	SMPL
N5844-02MSD - 1	MH-17MSD	15092	3.5321	-	7470A	PPB	05 Dec 2022 15:09:49	U	SMPL
CCV22 - 1	CCV22	20810	4.8472	-	7470A	PPB	05 Dec 2022 15:12:04	U	SMPL
CCB22 - 1	CCB22	32	0.0683	-	7470A	PPB	05 Dec 2022 15:14:19	U	SMPL
PB149405BL - 1	PBW	72	0.0775	-	7470A	PPB	05 Dec 2022 15:16:34	U	SMPL
PB149405BS - 1	LCSW	14531	3.4030	-	7470A	PPB	05 Dec 2022 15:18:49	U	SMPL
N5793-04 - 1	EB-2022-11-28	-101	0.0377	-	7470A	PPB	05 Dec 2022 15:21:05	U	SMPL
N5815-01 - 1	MW-01	-41	0.0515	-	7470A	PPB	05 Dec 2022 15:23:20	U	SMPL
N5815-03 - 1	MW-02	-22	0.0558	-	7470A	PPB	05 Dec 2022 15:25:36	U	SMPL
N5815-05 - 1	MW-03	32	0.0683	-	7470A	PPB	05 Dec 2022 15:27:52	U	SMPL
N5815-07 - 1	MW-04	-19	0.0565	-	7470A	PPB	05 Dec 2022 15:30:09	U	SMPL
N5815-09 - 1	MW-05	-13	0.0579	-	7470A	PPB	05 Dec 2022 15:32:25	U	SMPL
N5815-11 - 1	MW-08	-4	0.0600	-	7470A	PPB	05 Dec 2022 15:34:43	U	SMPL
N5815-13 - 1	MW-10	44	0.0710	-	7470A	PPB	05 Dec 2022 15:36:59	U	SMPL
CCV23 - 1	CCV23	20644	4.8090	-	7470A	PPB	05 Dec 2022 15:39:14	U	SMPL
CCB23 - 1	CCB23	-9	0.0588	-	7470A	PPB	05 Dec 2022 15:41:30	U	SMPL
N5815-13DUP - 1	MW-10DUP	84	0.0802	-	7470A	PPB	05 Dec 2022 15:43:46	U	SMPL
N5815-15 - 1	N5815-13MS	14526	3.4019	-	7470A	PPB	05 Dec 2022 15:48:37	U	SMPL
N5815-16 - 1	N5815-13MSD	15000	3.5109	-	7470A	PPB	05 Dec 2022 15:50:53	U	SMPL
N5815-19 - 1	PZ-2	489	0.1734	-	7470A	PPB	05 Dec 2022 15:53:08	U	SMPL
N5815-21 - 1	PZ-3R	315	0.1333	-	7470A	PPB	05 Dec 2022 15:55:25	U	SMPL
N5815-23 - 1	RW-1	-14	0.0577	-	7470A	PPB	05 Dec 2022 15:57:41	U	SMPL
N5815-25 - 1	MW-15	31	0.0680	-	7470A	PPB	05 Dec 2022 15:59:57	U	SMPL
N5830-01 - 1	COMP1-1201	205	0.1080	-	7470A	PPB	05 Dec 2022 16:02:14	U	SMPL
N5847-01 - 1	B-40-GW-01	924	0.2734	-	7470A	PPB	05 Dec 2022 16:04:31	U	SMPL
N5852-02 - 1	RW9-PT-IN-03-12012022	-24	0.0554	-	7470A	PPB	05 Dec 2022 16:06:47	U	SMPL
CCV24 - 1	CCV24	22371	5.2063	-	7470A	PPB	05 Dec 2022 16:09:02	U	SMPL
CCB24 - 1	CCB24	-81	0.0423	-	7470A	PPB	05 Dec 2022 16:11:17	U	SMPL
N5852-03 - 1	RW9-PT-OUT-03-12012022	35	0.0689	-	7470A	PPB	05 Dec 2022 16:13:33	U	SMPL
PB149406BL - 1	PBW	59	0.0745	-	7470A	PPB	05 Dec 2022 16:15:48	U	SMPL
PB149406BS - 1	LCSW	16668	3.8946	-	7470A	PPB	05 Dec 2022 16:20:56	U	SMPL
N5815-02 - 1	MW-01	-34	0.0531	-	7470A	PPB	05 Dec 2022 16:23:12	U	SMPL
N5815-04 - 1	MW-02	48	0.0719	-	7470A	PPB	05 Dec 2022 16:25:27	U	SMPL
N5815-06 - 1	MW-03	77	0.0786	-	7470A	PPB	05 Dec 2022 16:27:44	U	SMPL
N5815-08 - 1	MW-04	0	0.0609	-	7470A	PPB	05 Dec 2022 16:30:01	U	SMPL
N5815-10 - 1	MW-05	-12	0.0581	-	7470A	PPB	05 Dec 2022 16:32:18	U	SMPL
N5815-12 - 1	MW-08	15	0.0643	-	7470A	PPB	05 Dec 2022 16:34:35	U	SMPL
N5815-14 - 1	MW-10	8	0.0627	-	7470A	PPB	05 Dec 2022 16:36:51	U	SMPL
CCV25 - 1	CCV25	23381	5.4386	-	7470A	PPB	05 Dec 2022 16:39:07	U	SMPL
CCB25 - 1	CCB25	-90	0.0402	-	7470A	PPB	05 Dec 2022 16:41:23	U	SMPL
N5815-14DUP - 1	MW-10DUP	71	0.0772	-	7470A	PPB	05 Dec 2022 16:43:39	U	SMPL
N5815-17 - 1	N5815-14MS	16982	3.9668	-	7470A	PPB	05 Dec 2022 17:02:47	U	SMPL
N5815-18 - 1	N5815-14MSD	17249	4.0282	-	7470A	PPB	05 Dec 2022 17:06:59	U	SMPL
N5815-20 - 1	PZ-2	276	0.1244	-	7470A	PPB	05 Dec 2022 17:09:14	U	SMPL
N5815-22 - 1	PZ-3R	373	0.1467	-	7470A	PPB	05 Dec 2022 17:11:34	U	SMPL
N5815-24 - 1	RW-1	55	0.0735	-	7470A	PPB	05 Dec 2022 17:13:51	U	SMPL
N5815-26 - 1	MW-15	110	0.0862	-	7470A	PPB	05 Dec 2022 17:16:12	U	SMPL
PB149376TB - 1	PB149376TB	59	0.0745	-	7470A	PPB	05 Dec 2022 17:18:29	U	SMPL
CCV26 - 1	CCV26	21083	4.9100	-	7470A	PPB	05 Dec 2022 17:23:21	U	SMPL
CCB26 - 1	CCB26	129	0.0906	-	7470A	PPB	05 Dec 2022 17:27:35	U	SMPL
N5844-02LX5 - 1		67	0.0763	-	7470A	PPB	05 Dec 2022 17:29:51	U	SMPL
N5844-02A - 1		16952	3.9599	-	7470A	PPB	05 Dec 2022 17:32:07	U	SMPL
N5815-13LX5 - 1		-61	0.0469	-	7470A	PPB	05 Dec 2022 17:34:23	U	SMPL
N5815-13A - 1		17013	3.9739	-	7470A	PPB	05 Dec 2022 17:36:39	U	SMPL
N5815-14LX5 - 1		-32	0.0535	-	7470A	PPB	05 Dec 2022 17:38:56	U	SMPL
N5815-14A - 1		17043	3.9808	-	7470A	PPB	05 Dec 2022 17:41:12	U	SMPL
CCV27 - 1	CCV27	23041	5.3604	-	7470A	PPB	05 Dec 2022 17:43:28	U	SMPL
CCB27 - 1	CCB27	-23	0.0556	-	7470A	PPB	05 Dec 2022 17:45:45	U	SMPL