



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123112

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	12/05/2022	11:40
2	N5850-01		1	11	100.00	>212.0	>212.0	12/05/2022	12:10
3	N5850-01DUP		1	14	100.00	>212.0	>212.0	12/05/2022	12:40
4	N5859-03		1	12	100.00	>212.0	>212.0	12/05/2022	13:10
5	N5859-04		1	14	100.00	>212.0	>212.0	12/05/2022	13:40
6	N5859-05		1	13	100.00	>212.0	>212.0	12/05/2022	14:10
7	N5859-06		1	13	100.00	>212.0	>212.0	12/05/2022	14:40
8	N5859-07		1	14	100.00	>212.0	>212.0	12/05/2022	15:10
9	N5859-08		1	12	100.00	>212.0	>212.0	12/05/2022	15:40
10	N5859-09		1	14	100.00	>212.0	>212.0	12/05/2022	16:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66123112

WorkList Name : fp12-5

WorkList ID : 165480

Department : Wet-Chemistry

Date : 12-05-2022 08:49:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/08/2022	Water	N5850-01	Flash Point	Cool 4 deg C	PSEG03	M11	PURGE-WATER	12/02/2022	1010B
12/07/2022	Water	N5859-03	Flash Point	Cool 4 deg C	PSEG03	M11	ORA-1518	12/02/2022	1010B
12/07/2022	Water	N5859-04	Flash Point	Cool 4 deg C	PSEG03	M11	ORA-0583	12/02/2022	1010B
12/07/2022	Water	N5859-05	Flash Point	Cool 4 deg C	PSEG03	M11	ORA-0584	12/02/2022	1010B
12/07/2022	Water	N5859-06	Flash Point	Cool 4 deg C	PSEG03	M11	ORA-0585	12/02/2022	1010B
12/07/2022	Water	N5859-07	Flash Point	Cool 4 deg C	PSEG03	M11	ORA-0586	12/02/2022	1010B
12/07/2022	Water	N5859-08	Flash Point	Cool 4 deg C	PSEG03	M11	ORA-0587	12/02/2022	1010B
12/07/2022	Water	N5859-09	Flash Point	Cool 4 deg C	PSEG03	M11	31910	12/02/2022	1010B

Date/Time 12/15/2022 11:30
 Raw Sample Received by: RMA CWC
 Raw Sample Relinquished by: Co(SM)

Date/Time 12/15/2022 16:15
 Raw Sample Received by: Co(SM)
 Raw Sample Relinquished by: RMA (WC)