



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123113

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	12/06/2022	08:40
2	N5859-01		1	12	100.00	>212.0	>212.0	12/06/2022	09:10
3	N5859-01DUP		1	14	100.00	>212.0	>212.0	12/06/2022	09:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66123113

WorkList Name : fp-1

WorkList ID : 165492

Department : Wet-Chemistry

Date : 12-05-2022 12:20:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/07/2022	Solid	N5859-01	Flash Point	Cool 4 deg C	PSEG03	M11	ORA-0585-OIL	12/02/2022	1010B

Date/Time 12/6/2022 08.30
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: CP(SM)

Date/Time 12/6/2022 10.15
 Raw Sample Received by: CP(SM)
 Raw Sample Relinquished by: RM CWC