

LB123118


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

4609E



A= 0.0000e+000

B= 2.1570e-004

C= 4.2231e-002

Rho= 0.9998890

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/1/3
0.0	0.000	0.076	0.076	156	0.000	156					-7
0.2	0.200	0.187	-0.013	669	0.0 %	669					-4
2.5	2.500	2.404	-0.096	10949	0.0 %	10949					0
5.0	5.000	4.996	-0.004	22967	0.0 %	22967					1
7.5	7.500	7.552	0.052	34817	0.0 %	34817					6
10.0	10.000	9.985	-0.015	46096	0.0 %	46096					

LB123118 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Dec 2022 12:27:34

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	156	-	0.0000	7470A	PPB	06 Dec 2022 12:41:11	S	Std
0.2 - 1	50.2	669	-	0.2000	7470A	PPB	06 Dec 2022 12:45:27	S	Std
2.5 - 1	52.5	10949	-	2.5000	7470A	PPB	06 Dec 2022 12:47:44	S	Std
5.0 - 1	55	22967	-	5.0000	7470A	PPB	06 Dec 2022 12:50:01	S	Std
7.5 - 1	57.5	34817	-	7.5000	7470A	PPB	06 Dec 2022 12:54:51	S	Std
10.0 - 1	51.0	46096	-	10.0000	7470A	PPB	06 Dec 2022 12:59:43	S	Std
ICV46 - 1	ICV46	18484	4.0292	-	7470A	PPB	06 Dec 2022 13:05:18	U	SMPL
ICB46 - 1	ICB46	-75	0.0261	-	7470A	PPB	06 Dec 2022 13:07:33	U	SMPL
CCV28 - 1	CCV28	22757	4.9509	-	7470A	PPB	06 Dec 2022 13:09:48	U	SMPL
CCB28 - 1	CCB28	-157	0.0084	-	7470A	PPB	06 Dec 2022 13:12:03	U	SMPL
CRA - 1	CRA	559	0.1628	-	7470A	PPB	06 Dec 2022 13:14:18	U	SMPL
HighStd - 1	HighStd	44732	9.6909	-	7470A	PPB	06 Dec 2022 13:16:34	U	SMPL
ChkStd - 1	ChkStd	31739	6.8883	-	7470A	PPB	06 Dec 2022 13:18:49	U	SMPL
PB149443BL - 1	PBW	-39	0.0338	-	7470A	PPB	06 Dec 2022 13:21:04	U	SMPL
PB149443BS - 1	LCSW	15786	3.4477	-	7470A	PPB	06 Dec 2022 13:23:20	U	SMPL
N5850-01 - 1	PURGE-WATER	548	0.1604	-	7470A	PPB	06 Dec 2022 13:37:15	U	SMPL
N5850-01DUP - 1	PURGE-WATERDUP	480	0.1458	-	7470A	PPB	06 Dec 2022 13:39:31	U	SMPL
N5850-01MS - 1	PURGE-WATERMS	18139	3.9548	-	7470A	PPB	06 Dec 2022 13:41:47	U	SMPL
N5850-01MSD - 1	PURGE-WATERMSD	18116	3.9499	-	7470A	PPB	06 Dec 2022 13:44:04	U	SMPL
N5857-01 - 1	COMP-SAMPLE-TANK-10-11-12-31-33	609	0.1736	-	7470A	PPB	06 Dec 2022 13:46:20	U	SMPL
CCV29 - 1	CCV29	22710	4.9408	-	7470A	PPB	06 Dec 2022 13:48:35	U	SMPL
CCB29 - 1	CCB29	-50	0.0314	-	7470A	PPB	06 Dec 2022 13:50:50	U	SMPL
N5858-01 - 1	COMP-SAMPLE-26-27-28-30	639	0.1801	-	7470A	PPB	06 Dec 2022 13:53:05	U	SMPL
N5859-03 - 1	ORA-1518	666	0.1859	-	7470A	PPB	06 Dec 2022 13:55:21	U	SMPL
N5859-06 - 1	ORA-0585	408	0.1302	-	7470A	PPB	06 Dec 2022 13:57:36	U	SMPL
N5859-08 - 1	ORA-0587	882	0.2328	-	7470A	PPB	06 Dec 2022 13:59:53	U	SMPL
N5882-01 - 1	FM1-114	598	0.1712	-	7470A	PPB	06 Dec 2022 14:02:09	U	SMPL
N5850-01LX5 - 1		-59	0.0295	-	7470A	PPB	06 Dec 2022 14:04:25	U	SMPL
N5850-01A - 1		17410	3.7976	-	7470A	PPB	06 Dec 2022 14:06:41	U	SMPL
CCV30 - 1	CCV30	21236	4.6228	-	7470A	PPB	06 Dec 2022 14:08:58	U	SMPL
CCB30 - 1	CCB30	-283	-0.0188	-	7470A	PPB	06 Dec 2022 14:11:17	U	SMPL