

LB123123


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

47539



A= 0.0000e+000

B= 2.0882e-004

C= 2.9921e-002

Rho= 0.9999319

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/d
0.0	0.000	0.024	0.024	-27	0.000	-27					-1
0.2	0.200	0.199	-0.001	811	0.0 %	811					-2
2.5	2.500	2.450	-0.050	11587	0.0 %	11587					0
5.0	5.000	4.992	-0.008	23761	0.0 %	23761					1
7.5	7.500	7.578	0.078	36146	0.0 %	36146					0
10.0	10.000	9.957	-0.043	47539	0.0 %	47539					

LB123123 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 Dec 2022 14:35:06

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-27	-	0.0000	7471B	PPB	06 Dec 2022 14:39:09	S	Std
0.2 - 1	50.2	811	-	0.2000	7471B	PPB	06 Dec 2022 14:41:26	S	Std
2.5 - 1	52.5	11587	-	2.5000	7471B	PPB	06 Dec 2022 14:43:43	S	Std
5.0 - 1	55	23761	-	5.0000	7471B	PPB	06 Dec 2022 14:46:00	S	Std
7.5 - 1	57.5	36146	-	7.5000	7471B	PPB	06 Dec 2022 14:48:17	S	Std
10.0 - 1	51	47539	-	10.0000	7471B	PPB	06 Dec 2022 14:50:32	S	Std
ICV47 - 1	ICV47	20504	4.3116	-	7471B	PPB	06 Dec 2022 14:53:12	U	SMPL
ICB47 - 1	ICB47	16	0.0333	-	7471B	PPB	06 Dec 2022 14:55:27	U	SMPL
CCV31 - 1	CCV31	23572	4.9523	-	7471B	PPB	06 Dec 2022 14:57:42	U	SMPL
CCB31 - 1	CCB31	-73	0.0147	-	7471B	PPB	06 Dec 2022 14:59:57	U	SMPL
CRA - 1	CRA	810	0.1991	-	7471B	PPB	06 Dec 2022 15:02:12	U	SMPL
HighStd - 1	HighStd	47253	9.8974	-	7471B	PPB	06 Dec 2022 15:04:27	U	SMPL
ChkStd - 1	ChkStd	34039	7.1380	-	7471B	PPB	06 Dec 2022 15:06:43	U	SMPL
PB149442BL - 1	PBS	-88	0.0115	-	7471B	PPB	06 Dec 2022 15:09:00	U	SMPL
PB149442BS - 1	LCSS	17094	3.5995	-	7471B	PPB	06 Dec 2022 15:11:16	U	SMPL
N5849-01 - 1	TR-5-120222	3552	0.7717	-	7471B	PPB	06 Dec 2022 15:13:33	U	SMPL
N5859-01 - 1	ORA-0585-OIL	1263	0.2937	-	7471B	PPB	06 Dec 2022 15:15:49	U	SMPL
N5863-01 - 1	SAND-PILE-1	1800	0.4058	-	7471B	PPB	06 Dec 2022 15:18:06	U	SMPL
N5863-01DUP - 1	SAND-PILE-1DUP	1625	0.3693	-	7471B	PPB	06 Dec 2022 15:20:24	U	SMPL
N5863-01MS - 1	SAND-PILE-1MS	19221	4.0437	-	7471B	PPB	06 Dec 2022 15:22:39	U	SMPL
CCV32 - 1	CCV32	23455	4.9279	-	7471B	PPB	06 Dec 2022 15:24:55	U	SMPL
CCB32 - 1	CCB32	-83	0.0126	-	7471B	PPB	06 Dec 2022 15:27:10	U	SMPL
N5863-01MSD - 1	SAND-PILE-1MSD	19548	4.1120	-	7471B	PPB	06 Dec 2022 15:29:25	U	SMPL
N5863-02 - 1	WOOD-CHIP-PILE	3539	0.7689	-	7471B	PPB	06 Dec 2022 15:31:41	U	SMPL
N5868-01 - 1	TP-0	29655	6.2225	-	7471B	PPB	06 Dec 2022 15:33:56	U	SMPL
N5870-01 - 1	ML-1	640	0.1636	-	7471B	PPB	06 Dec 2022 15:36:12	U	SMPL
N5870-02 - 1	ML-2	751	0.1867	-	7471B	PPB	06 Dec 2022 15:38:28	U	SMPL
N5870-03 - 1	ML-3	1222	0.2851	-	7471B	PPB	06 Dec 2022 15:40:44	U	SMPL
N5870-04 - 1	ML-4	332	0.0993	-	7471B	PPB	06 Dec 2022 15:43:01	U	SMPL
N5870-05 - 1	ML-5	612	0.1577	-	7471B	PPB	06 Dec 2022 15:45:18	U	SMPL
N5871-01 - 1	SU-4-120222	1558	0.3553	-	7471B	PPB	06 Dec 2022 15:47:35	U	SMPL
N5872-01 - 1	HR-02-120222	512	0.1368	-	7471B	PPB	06 Dec 2022 15:49:50	U	SMPL
CCV33 - 1	CCV33	23189	4.8723	-	7471B	PPB	06 Dec 2022 15:52:06	U	SMPL
CCB33 - 1	CCB33	-64	0.0166	-	7471B	PPB	06 Dec 2022 15:54:22	U	SMPL
N5881-01 - 1	VNJ-202	4594	0.9893	-	7471B	PPB	06 Dec 2022 15:56:37	U	SMPL
N5883-01 - 1	MH-18	266	0.0855	-	7471B	PPB	06 Dec 2022 15:58:53	U	SMPL
N5884-01 - 1	EO-3-120522	1055	0.2502	-	7471B	PPB	06 Dec 2022 16:01:08	U	SMPL
N5885-01 - 1	JC-04-120522	36898	7.7350	-	7471B	PPB	06 Dec 2022 16:03:24	U	SMPL
N5863-01LX5 - 1		188	0.0692	-	7471B	PPB	06 Dec 2022 16:05:41	U	SMPL
N5863-01A - 1		20218	4.2519	-	7471B	PPB	06 Dec 2022 16:07:57	U	SMPL
CCV34 - 1	CCV34	23406	4.9176	-	7471B	PPB	06 Dec 2022 16:10:13	U	SMPL
CCB34 - 1	CCB34	-29	0.0239	-	7471B	PPB	06 Dec 2022 16:12:31	U	SMPL