



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB123124

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	12/06/2022	14:50
2	CAL2	1	Water	NA	NA	20.7	7.01	12/06/2022	14:51
3	CAL3	1	Water	NA	NA	20.7	10.03	12/06/2022	14:52
4	ICV	1	Water	NA	NA	20.7	6.98	12/06/2022	15:00
5	CCV1	1	Water	NA	NA	20.7	2.01	12/06/2022	15:02
6	N5882-01	1	Water	NA	NA	20.6	8.01	12/06/2022	15:05
7	N5890-03	1	Water	NA	NA	20.1	6.92	12/06/2022	15:10
8	N5890-05	1	Water	NA	NA	20.1	6.88	12/06/2022	15:11
9	N5890-06	1	Water	NA	NA	20.2	6.90	12/06/2022	15:15
10	N5890-07	1	Water	NA	NA	20.2	10.94	12/06/2022	15:20
11	N5890-09	1	Water	NA	NA	20.7	10.99	12/06/2022	15:25
12	N5890-10	1	Water	NA	NA	20.2	11.04	12/06/2022	15:30
13	N5890-11	1	Water	NA	NA	20.1	11.00	12/06/2022	15:31
14	N5890-11DUP	1	Water	NA	NA	20.2	11.01	12/06/2022	15:33
15	CCV2	1	Water	NA	NA	20.7	12.03	12/06/2022	15:35

213123124

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n5889 WorkList ID : 165528 Department : Wet-Chemistry Date : 12-06-2022 14:32:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/05/2022	Water	N5882-01	pH	Cool 4 deg C	PSEG03	M11	FM1-114	12/05/2022	9040C
12/06/2022	Water	N5890-03	pH	Cool 4 deg C	PSEG03	M11	COMP-1	12/06/2022	9040C
12/06/2022	Water	N5890-05	pH	Cool 4 deg C	PSEG03	M11	1826	12/06/2022	9040C
12/06/2022	Water	N5890-06	pH	Cool 4 deg C	PSEG03	M11	1827	12/06/2022	9040C
12/06/2022	Water	N5890-07	pH	Cool 4 deg C	PSEG03	M11	COMP-2	12/06/2022	9040C
12/06/2022	Water	N5890-09	pH	Cool 4 deg C	PSEG03	M11	1920	12/06/2022	9040C
12/06/2022	Water	N5890-10	pH	Cool 4 deg C	PSEG03	M11	1921	12/06/2022	9040C
12/06/2022	Water	N5890-11	pH	Cool 4 deg C	PSEG03	M11	1922	12/06/2022	9040C

Date/Time 12/06/22 1440

Raw Sample Received by: 701001

Raw Sample Relinquished by: [Signature]

Date/Time 12/06/22 16120

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: 701001