



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/7/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 12/06/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:29
Out Date: 12/07/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB123129

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N5896-01	DBXL9	1	1.15	8.77	9.92	9.52	95.4	
N5896-02	DBXM0	2	1.17	8.81	9.98	8.81	86.7	
N5896-03	DBXM1	3	1.14	8.83	9.97	9.44	94.0	
N5896-04	DBXM2	4	1.13	8.72	9.85	9.44	95.3	
N5896-05	DBXM3	5	1.14	8.50	9.64	9.00	92.5	
N5896-06	DBXM4	6	1.14	8.38	9.52	8.91	92.7	
N5896-07	DBXM5	7	1.12	8.86	9.98	8.82	86.9	
N5896-08	DBXM6	8	1.14	8.76	9.9	7.65	74.3	
N5896-09	DBXM7	9	1.15	8.82	9.97	8.13	79.1	
N5896-10	DBXM8	10	1.13	8.50	9.63	8.78	90.0	
N5896-11	DBXM9	11	1.13	8.72	9.85	9.19	92.4	
N5896-12	DBXN0	12	1.14	8.70	9.84	9.2	92.6	
N5896-13	DBXN1	13	1.14	8.66	9.8	9.11	92.0	
N5896-14	DBXN2	14	1.14	8.84	9.98	9.25	91.7	
N5896-15	DBXN3	15	1.13	8.62	9.75	8.65	87.2	
N5896-16	DBXN4	16	1.13	8.84	9.97	8.88	87.7	
N5896-17	DBXN5	17	1.12	8.78	9.9	8.61	85.3	
N5896-18	DBXN6	18	1.14	8.84	9.98	8.92	88.0	
N5896-19	DBXN7	19	1.13	8.83	9.96	9.04	89.6	
N5896-20	DBXN8	20	1.13	8.85	9.98	8.66	85.1	
N5896-21	DBXN8MS	21	1.13	8.85	9.98	8.66	85.1	
N5896-22	DBXN8MSD	22	1.13	8.85	9.98	8.66	85.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$